

LB 112303

SEAM01.0

INSTRUMENT ID: CV1


SFAM01.1

MB
2/24/20

Linear

μ Abs.:

42675



A= 0.0000e+000

B= 2.3487e-004

C= -2.9262e-002

Rho= 0.9999846

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.016	-0.016	55	0.000	55					-
0.05	0.050					244					-
0.20	0.200	0.184	-0.016	907	0.0 %	907					-8
2.50	2.500	2.536	0.036	10923	0.0 %	10923					1
5.00	5.000	5.019	0.019	21493	0.0 %	21493					0
7.50	7.500	7.483	-0.017	31986	0.0 %	31986					0
10.0	10.000	9.994	-0.006	42675	0.0 %	42675					0

LB112303
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	55	-		0 SFAM01.0	PPB	12/10/2020 14:03	Std
	0.2 S0.2	907	-		0.2 SFAM01.0	PPB	12/10/2020 14:06	Std
	2.5 S2.5	10923	-		2.5 SFAM01.0	PPB	12/10/2020 14:08	Std
	5 S5.0	21493	-		5 SFAM01.0	PPB	12/10/2020 14:10	Std
	7.5 S7.5	31986	-		7.5 SFAM01.0	PPB	12/10/2020 14:13	Std
	10 S10.0	42675	-		10 SFAM01.0	PPB	12/10/2020 14:15	Std
ICV	ICV	16512	3.849 -		SFAM01.0	PPB	12/10/2020 14:18	SMPL
ICB	ICB	-39	-0.0384 -		SFAM01.0	PPB	12/10/2020 14:20	SMPL
CCV055	CCV055	21209	4.9522 -		SFAM01.0	PPB	12/10/2020 14:22	SMPL
CCB055	CCB055	-43	-0.0394 -		SFAM01.0	PPB	12/10/2020 14:24	SMPL
PB133528BL	PBW528	21	-0.0243 -		SFAM01.0	PPB	12/10/2020 14:27	SMPL
L4956-07	X2R70	10545	2.4475 -		SFAM01.0	PPB	12/10/2020 14:29	SMPL
L4956-08	X2R70D	10537	2.4456 -		SFAM01.0	PPB	12/10/2020 14:33	SMPL
L4956-09	X2R70S	15152	3.5296 -		SFAM01.0	PPB	12/10/2020 14:38	SMPL
CCV056	CCV056	21131	4.9339 -		SFAM01.0	PPB	12/10/2020 14:40	SMPL
CCB056	CCB056	-46	-0.0401 -		SFAM01.0	PPB	12/10/2020 14:42	SMPL