



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112306

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	12/10/2020	09:45
2	L4985-05		1	8	100.00	>212.0	>212.0	12/10/2020	10:10
3	L4985-05DUP		1	8	100.00	>212.0	>212.0	12/10/2020	10:40
4	L4986-03		1	9	100.00	>212.0	>212.0	12/10/2020	11:15
5	L4987-01		1	9	100.00	>212.0	>212.0	12/10/2020	11:50
6	L4987-03		1	10	100.00	>212.0	>212.0	12/10/2020	12:20
7	L4987-05		1	9	100.00	>212.0	>212.0	12/10/2020	12:50
8	L4988-03		1	11	100.00	>212.0	>212.0	12/10/2020	13:20
9	L4991-01		1	10	100.00	>212.0	>212.0	12/10/2020	13:45
10	L4991-07		1	9	100.00	>212.0	>212.0	12/10/2020	14:15
11	L4992-01		1	11	100.00	>212.0	>212.0	12/10/2020	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 112306

WorkList Name : fp121020

WorkList ID : 144800

Department : Wet-Chemistry

Date : 12-10-2020 09:00:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2020	Water	L4985-05 <i>C</i>	Flash Point	Cool 4 deg C	PSEG03	L21	32831	12/09/2020	1010A
12/16/2020	Water	L4986-03 <i>C</i>	Flash Point	Cool 4 deg C	PSEG03	L21	32794	12/09/2020	1010A
12/16/2020	Water	L4987-01 <i>C</i>	Flash Point	Cool 4 deg C	PSEG03	L21	32833	12/09/2020	1010A
12/16/2020	Water	L4987-03 <i>B</i>	Flash Point	Cool 4 deg C	PSEG03	L21	32832	12/09/2020	1010A
12/16/2020	Water	L4987-05 <i>E</i>	Flash Point	Cool 4 deg C	PSEG03	L21	32834	12/09/2020	1010A
12/16/2020	Water	L4988-03 <i>C</i>	Flash Point	Cool 4 deg C	PSEG03	L22	32818	12/09/2020	1010A
12/22/2020	Water	L4991-01 <i>I</i>	Flash Point	Cool 4 deg C	PSEG03	L21	COMP-2	12/09/2020	1010A
12/22/2020	Water	L4991-07 <i>C</i>	Flash Point	Cool 4 deg C	PSEG03	L21	COMP-1	12/09/2020	1010A
12/22/2020	Water	L4992-01 <i>I</i>	Flash Point	Cool 4 deg C	PSEG03	L22	TRE-1130	12/09/2020	1010A

Date/Time 12/10/20 09:15
 Received by: *KS (wetchem)*
 Relinquished by: *AP (Ext. Lab)*

Date/Time 12/10/20 14:55
 Received by: *AP (Ext. Lab)*
 Relinquished by: *KS (wetchem)*