


LB112308

INSTRUMENT ID: CV1

7471B

μ Abs.:

53091



Linear

A= 0.0000e+000

B= 1.8893e-004

C= 1.2394e-002

Rho= 0.9998403

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	41	0.000	41				
0.2	0.200	0.197	-0.003	978	0.0 %	978				
2.5	2.500	2.452	-0.048	12912	0.0 %	12912				
5.0	5.000	5.097	0.097	26915	0.0 %	26915				
7.5	7.500	7.390	-0.110	39051	0.0 %	39051				
10.0	10.000	10.043	0.043	53091	0.0 %	53091				

LB112308 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 10 Dec 2020 15:44:09					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7471B	PPB	10 Dec 2020 15:47:10	S	Std
0.2 - 1	50.2	978	-	0.2000	7471B	PPB	10 Dec 2020 15:49:28	S	Std
2.5 - 1	52.5	12912	-	2.5000	7471B	PPB	10 Dec 2020 15:51:45	S	Std
5.0 - 1	55	26915	-	5.0000	7471B	PPB	10 Dec 2020 15:54:03	S	Std
7.5 - 1	57.5	39051	-	7.5000	7471B	PPB	10 Dec 2020 15:56:19	S	Std
10.0 - 1	510	53091	-	10.0000	7471B	PPB	10 Dec 2020 16:01:08	S	Std
ICV36 - 1	ICV36	20349	3.8570	-	7471B	PPB	10 Dec 2020 16:03:48	U	SMPL
ICB36 - 1	ICB36	-73	-0.0014	-	7471B	PPB	10 Dec 2020 16:06:02	U	SMPL
CCV57 - 1	CCV57	27163	5.1443	-	7471B	PPB	10 Dec 2020 16:08:19	U	SMPL
CCB57 - 1	CCB57	-78	-0.0023	-	7471B	PPB	10 Dec 2020 16:10:35	U	SMPL
CRA - 1	CRA	946	0.1911	-	7471B	PPB	10 Dec 2020 16:12:49	U	SMPL
HighStd - 1	HighStd	53938	10.2030	-	7471B	PPB	10 Dec 2020 16:15:04	U	SMPL
ChkStd - 1	ChkStd	37436	7.0852	-	7471B	PPB	10 Dec 2020 16:17:20	U	SMPL
PB133521BL - 1	PBS	-93	-0.0052	-	7471B	PPB	10 Dec 2020 16:19:37	U	SMPL
PB133521BS - 1	LCSS	19219	3.6435	-	7471B	PPB	10 Dec 2020 16:21:53	U	SMPL
L4989-01 - 1	BU-02-120920	2240	0.4356	-	7471B	PPB	10 Dec 2020 16:24:09	U	SMPL
L5008-01 - 1	S-1	2125	0.4139	-	7471B	PPB	10 Dec 2020 16:26:26	U	SMPL
L5008-01DUP - 1	S-1DUP	1801	0.3527	-	7471B	PPB	10 Dec 2020 16:28:43	U	SMPL
L5008-01MS - 1	S-1MS	21718	4.1156	-	7471B	PPB	10 Dec 2020 16:30:59	U	SMPL
L5008-01MSD - 1	S-1MSD	21395	4.0546	-	7471B	PPB	10 Dec 2020 16:33:15	U	SMPL
CCV58 - 1	CCV58	27092	5.1309	-	7471B	PPB	10 Dec 2020 16:35:30	U	SMPL
CCB58 - 1	CCB58	-65	0.0001	-	7471B	PPB	10 Dec 2020 16:37:46	U	SMPL
L5009-02 - 1	KENS-B5-120920-0-8	2100	0.4092	-	7471B	PPB	10 Dec 2020 16:40:02	U	SMPL
L5009-03 - 1	KENS-B5-120920-13-15	2830	0.5471	-	7471B	PPB	10 Dec 2020 16:42:18	U	SMPL
L5009-06 - 1	KENS-B5-120920-DP	1333	0.2642	-	7471B	PPB	10 Dec 2020 16:44:33	U	SMPL
L5008-01LX5 - 1		338	0.0763	-	7471B	PPB	10 Dec 2020 16:46:49	U	SMPL
L5008-01A - 1		23373	4.4283	-	7471B	PPB	10 Dec 2020 16:49:05	U	SMPL
CCV59 - 1	CCV59	26720	5.0606	-	7471B	PPB	10 Dec 2020 16:51:21	U	SMPL
CCB59 - 1	CCB59	-67	-0.0003	-	7471B	PPB	10 Dec 2020 16:53:38	U	SMPL