


LB112318

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.040	0.040	39	0.000	39				
0.2	0.200	0.228	0.028	721	0.0 %	721				
2.5	2.500	2.471	-0.029	8879	0.0 %	8879				
5.0	5.000	4.902	-0.098	17719	0.0 %	17719				
7.5	7.500	7.516	0.016	27225	0.0 %	27225				
10.0	10.000	10.042	0.042	36408	0.0 %	36408				

LB112318 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Dec 2020 10:48:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7470A	PPB	11 Dec 2020 10:59:05	S	Std
0.2 - 1	50.2	721	-	0.2000	7470A	PPB	11 Dec 2020 11:01:23	S	Std
2.5 - 1	52.5	8879	-	2.5000	7470A	PPB	11 Dec 2020 11:03:40	S	Std
5.0 - 1	55	17719	-	5.0000	7470A	PPB	11 Dec 2020 11:05:57	S	Std
7.5 - 1	57.5	27225	-	7.5000	7470A	PPB	11 Dec 2020 11:08:13	S	Std
10.0 - 1	510	36408	-	10.0000	7470A	PPB	11 Dec 2020 11:10:29	S	Std
ICV37 - 1	ICV37	13466	3.7327	-	7470A	PPB	11 Dec 2020 11:13:18	U	SMPL
ICB37 - 1	ICB37	-57	0.0140	-	7470A	PPB	11 Dec 2020 11:15:33	U	SMPL
CCV60 - 1	CCV60	17891	4.9496	-	7470A	PPB	11 Dec 2020 11:17:47	U	SMPL
CCB60 - 1	CCB60	-53	0.0151	-	7470A	PPB	11 Dec 2020 11:20:03	U	SMPL
CRA - 1	CRA	672	0.2144	-	7470A	PPB	11 Dec 2020 11:22:18	U	SMPL
HighStd - 1	HighStd	36459	10.0557	-	7470A	PPB	11 Dec 2020 11:24:33	U	SMPL
ChkStd - 1	ChkStd	25784	7.1201	-	7470A	PPB	11 Dec 2020 11:26:48	U	SMPL
PB133526BL - 1	PBW	-95	0.0035	-	7470A	PPB	11 Dec 2020 11:29:05	U	SMPL
PB133526BS - 1	LCSW	13798	3.8240	-	7470A	PPB	11 Dec 2020 11:31:21	U	SMPL
L4991-01 - 1	COMP-2	391	0.1372	-	7470A	PPB	11 Dec 2020 11:39:38	U	SMPL
L4991-01DUP - 1	COMP-2DUP	233	0.0937	-	7470A	PPB	11 Dec 2020 11:41:54	U	SMPL
L4991-01MS - 1	COMP-2MS	12039	3.3403	-	7470A	PPB	11 Dec 2020 11:44:11	U	SMPL
L4991-01MSD - 1	COMP-2MSD	12221	3.3904	-	7470A	PPB	11 Dec 2020 11:46:28	U	SMPL
L4992-01 - 1	TRE-1130	0	0.0296	-	7470A	PPB	11 Dec 2020 11:48:44	U	SMPL
CCV61 - 1	CCV61	18419	5.0948	-	7470A	PPB	11 Dec 2020 11:51:01	U	SMPL
CCB61 - 1	CCB61	-25	0.0228	-	7470A	PPB	11 Dec 2020 11:53:16	U	SMPL
L5009-07 - 1	FIELD-BLANK	22	0.0357	-	7470A	PPB	11 Dec 2020 11:55:31	U	SMPL
L4991-01LX5 - 1		152	0.0714	-	7470A	PPB	11 Dec 2020 11:57:46	U	SMPL
L4991-01A - 1		13955	3.8672	-	7470A	PPB	11 Dec 2020 12:00:02	U	SMPL
CCV62 - 1	CCV62	18360	5.0786	-	7470A	PPB	11 Dec 2020 12:02:18	U	SMPL
CCB62 - 1	CCB62	-12	0.0263	-	7470A	PPB	11 Dec 2020 12:04:33	U	SMPL