



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112320

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	12/11/2020	13:10
2	CAL2	1	Water	NA	NA	21.4	7.00	12/11/2020	13:11
3	CAL3	1	Water	NA	NA	21.4	10.03	12/11/2020	13:12
4	ICV	1	Water	NA	NA	21.4	7.00	12/11/2020	13:15
5	CCV1	1	Water	NA	NA	21.4	2.02	12/11/2020	13:16
6	L4913-02	1	Solid	20.03	20	21.9	10.55	12/11/2020	13:18
7	L5022-01	1	Solid	20.02	20	21.4	9.71	12/11/2020	13:20
8	L5022-01DUP	1	Solid	20.03	20	21.5	9.72	12/11/2020	13:22
9	CCV2	1	Water	NA	NA	21.3	12.02	12/11/2020	13:23

112320

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4985 WorkList ID : 144813 Department : Wet-Chemistry Date : 12-11-2020 12:58:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/01/2020	Solid	L4913-02	Corrosivity	Cool 4 deg C	TULL02		TEST-PIT-3	12/01/2020	9045D
12/09/2020	Solid	L5022-01	Corrosivity	Cool 4 deg C	ENTA05	L51	WCS-OH-E6	12/09/2020	9045D

Date/Time 12-11-2020 13:00
Received by: SD Fleet Admin
Relinquished by: Jd

Date/Time 12-11-2020 14:00
Received by: Jd
Relinquished by: Jd Fleet Admin