



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112321

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	12/11/2020	13:10
2	CAL2	1	Water	NA	NA	21.4	7.00	12/11/2020	13:11
3	CAL3	1	Water	NA	NA	21.4	10.03	12/11/2020	13:12
4	ICV	1	Water	NA	NA	21.4	7.00	12/11/2020	13:15
5	CCV1	1	Water	NA	NA	21.4	2.02	12/11/2020	13:16
6	L5016-01	1	Solid	20.03	20	21.6	8.47	12/11/2020	13:25
7	L5016-01DUP	1	Solid	20.04	20	21.7	8.48	12/11/2020	13:26
8	CCV2	1	Water	NA	NA	21.4	12.02	12/11/2020	13:27

18112321

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4939 WorkList ID : 144864 Department : Wet-Chemistry Date : 12-11-2020 13:05:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2020	Solid	L5016-01	pH	Cool 4 deg C	PSEG05	L22	AU-06-121020	12/10/2020	9045D

Date/Time 12-11-2020 13:20
Received by: JO [signature]
Relinquished by: JO [signature]

Date/Time 12-11-2020 15:40
Received by: JO [signature]
Relinquished by: JO [signature]