



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112322

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	12/11/2020	13:10
2	CAL2	1	Water	NA	NA	21.4	7.00	12/11/2020	13:11
3	CAL3	1	Water	NA	NA	21.4	10.03	12/11/2020	13:12
4	ICV	1	Water	NA	NA	21.4	7.00	12/11/2020	13:15
5	CCV1	1	Water	NA	NA	21.4	2.02	12/11/2020	13:16
6	L5039-01	1	Water	NA	NA	18.3	7.89	12/11/2020	13:35
7	L5039-01DUP	1	Water	NA	NA	18.4	7.90	12/11/2020	13:36
8	CCV2	1	Water	NA	NA	21.3	12.02	12/11/2020	13:38

112322

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4940 W WorkList ID : 144865 Department : Wet-Chemistry Date : 12-11-2020 13:06:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2020	Water	L5039-01	pH		PARS15	L51	JBD-WC-L-SB-02	12/10/2020	9040C

Date/Time 12.11.2020 13:10
Received by: JD
Relinquished by: JD

Date/Time 12.11.2020 14:20
Received by: JD
Relinquished by: JD