

LB112326

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AM

Instrument ID : Konelab

12/11/2020 13:19

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.028	0.0	0.080	
ICB1	1.347	0.0	0.000	
CCV1	234.797	0.0	0.206	
CCB1	1.676	0.0	0.001	
PB133548BL	1.704	0.0	0.001	
L5022-01	1.329	0.0	0.000	
L5022-01DUP	1.432	0.0	0.000	
CCV2	248.796	0.0	0.219	
CCB2	1.833	0.0	0.001	

N	9
Mean	64.993
SD	104.5797
CV%	160.91

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 11 11:10:07 2020

Fri Dec 11 13:17:36 2020

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	Reactive CN	P	1.3825	µg/l	12/11/2020 12:43:20
5.OPPBCN	A	Reactive CN	P	5.9554	µg/l	12/11/2020 12:43:21
10PPBCN	A	Reactive CN	P	10.3273	µg/l	12/11/2020 12:43:22
50PPBCN	A	Reactive CN	P	48.9535	µg/l	12/11/2020 12:43:23
100PPBCN	A	Reactive CN	P	99.0427	µg/l	12/11/2020 12:43:24
250PPBCN	A	Reactive CN	P	248.1172	µg/l	12/11/2020 12:43:25
500PPBCN	A	Reactive CN	P	501.2214	µg/l	12/11/2020 12:43:26
ICV1	S	Reactive CN	P	92.0284	µg/l	12/11/2020 13:12:46
ICB1	S	Reactive CN	P	1.3467	µg/l	12/11/2020 13:12:47
CCV1	S	Reactive CN	P	234.7967	µg/l	12/11/2020 13:12:50
CCB1	S	Reactive CN	P	1.6758	µg/l	12/11/2020 13:12:51
PB133548BL	S	Reactive CN	P	1.7038	µg/l	12/11/2020 13:12:53
L5022-01	S	Reactive CN	P	1.3288	µg/l	12/11/2020 13:12:55
L5022-01DUP	S	Reactive CN	P	1.4321	µg/l	12/11/2020 13:17:32
CCV2	S	Reactive CN	P	248.7956	µg/l	12/11/2020 13:17:34
CCB2	S	Reactive CN	P	1.833	µg/l	12/11/2020 13:17:36

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

12/11/2020 12:46


Test Total CN

Accepted 12/11/2020 12:46

Factor 1133
Bias -0.001

Coeff. of det. 0.999951

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.000	1.3825	0.0000	
2	5.0PPBCN	0.004	5.9554	5.0000	
3	10PPBCN	0.008	10.3273	10.0000	
4	50PPBCN	0.042	48.9535	50.0000	
5	100PPBCN	0.086	99.0427	100.0000	
6	250PPBCN	0.218	248.1172	250.0000	
7	500PPBCN	0.441	501.2214	500.0000	