

LB112332


7471B

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

53370



A= 0.0000e+000

B= 1.8647e-004

C= -8.0230e-003

Rho= 0.9998719

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	60	0.000	60				
0.2	0.200	0.200	0.000	1118	0.0 %	1118				
2.5	2.500	2.420	-0.080	13020	0.0 %	13020				
5.0	5.000	5.102	0.102	27403	0.0 %	27403				
7.5	7.500	7.531	0.031	40431	0.0 %	40431				
10.0	10.000	9.944	-0.056	53370	0.0 %	53370				

LB112332 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Dec 2020 15:05:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	60	-	0.0000	7471B	PPB	11 Dec 2020 15:06:33	S	Std
0.2 - 1	50.2	1118	-	0.2000	7471B	PPB	11 Dec 2020 15:08:50	S	Std
2.5 - 1	52.5	13020	-	2.5000	7471B	PPB	11 Dec 2020 15:11:08	S	Std
5.0 - 1	55	27403	-	5.0000	7471B	PPB	11 Dec 2020 15:13:25	S	Std
7.5 - 1	57.5	40431	-	7.5000	7471B	PPB	11 Dec 2020 15:15:42	S	Std
10.0 - 1	510	53370	-	10.0000	7471B	PPB	11 Dec 2020 15:20:34	S	Std
ICV38 - 1	ICV38	21382	3.9790	-7471B	PPB	PPB	11 Dec 2020 15:23:18	U	SMPL
ICB38 - 1	ICB38	-64	-0.0200	-7471B	PPB	PPB	11 Dec 2020 15:25:33	U	SMPL
CCV65 - 1	CCV65	27408	5.1027	-7471B	PPB	PPB	11 Dec 2020 15:27:48	U	SMPL
CCB65 - 1	CCB65	-70	-0.0211	-7471B	PPB	PPB	11 Dec 2020 15:30:03	U	SMPL
CRA - 1	CRA	1097	0.1965	-7471B	PPB	PPB	11 Dec 2020 15:32:19	U	SMPL
HighStd - 1	HighStd	53818	10.0273	-7471B	PPB	PPB	11 Dec 2020 15:34:33	U	SMPL
ChkStd - 1	ChkStd	37943	7.0671	-7471B	PPB	PPB	11 Dec 2020 15:36:49	U	SMPL
PB133543BL - 1	PBS	-82	-0.0233	-7471B	PPB	PPB	11 Dec 2020 15:39:06	U	SMPL
PB133543BS - 1	LCSS	18876	3.5117	-7471B	PPB	PPB	11 Dec 2020 15:41:23	U	SMPL
L5014-01 - 1	351	576	0.0994	-7471B	PPB	PPB	11 Dec 2020 15:43:40	U	SMPL
L5015-02 - 1	TR-05-121020	2069	0.3778	-7471B	PPB	PPB	11 Dec 2020 15:45:57	U	SMPL
L5016-01 - 1	AU-06-121020	7240	1.3420	-7471B	PPB	PPB	11 Dec 2020 15:48:13	U	SMPL
L5033-05 - 1	CS-47	288	0.0457	-7471B	PPB	PPB	11 Dec 2020 15:50:31	U	SMPL
CCV66 - 1	CCV66	27204	5.0647	-7471B	PPB	PPB	11 Dec 2020 15:52:46	U	SMPL
CCB66 - 1	CCB66	-39	-0.0153	-7471B	PPB	PPB	11 Dec 2020 15:55:02	U	SMPL
L5033-06 - 1	CS-48	465	0.0787	-7471B	PPB	PPB	11 Dec 2020 15:57:17	U	SMPL
L5033-06DUP - 1	CS-48DUP	223	0.0336	-7471B	PPB	PPB	11 Dec 2020 15:59:32	U	SMPL
L5033-06MS - 1	CS-48MS	19274	3.5860	-7471B	PPB	PPB	11 Dec 2020 16:01:47	U	SMPL
L5033-06MSD - 1	CS-48MSD	19226	3.5770	-7471B	PPB	PPB	11 Dec 2020 16:04:03	U	SMPL
L5033-06LX5 - 1		-9	-0.0097	-7471B	PPB	PPB	11 Dec 2020 16:06:19	U	SMPL
L5033-06A - 1		22429	4.1743	-7471B	PPB	PPB	11 Dec 2020 16:08:35	U	SMPL
CCV67 - 1	CCV67	27414	5.1038	-7471B	PPB	PPB	11 Dec 2020 16:10:51	U	SMPL
CCB67 - 1	CCB67	-48	-0.0170	-7471B	PPB	PPB	11 Dec 2020 16:13:07	U	SMPL