



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112334

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	12/11/2020	15:00
2	CAL2	1	Water	NA	NA	21.3	7.01	12/11/2020	15:01
3	CAL3	1	Water	NA	NA	21.3	10.03	12/11/2020	15:02
4	ICV	1	Water	NA	NA	21.3	7.00	12/11/2020	15:05
5	CCV1	1	Water	NA	NA	21.3	2.01	12/11/2020	15:06
6	L5070-01	1	Water	NA	NA	19.1	5.23	12/11/2020	15:10
7	L5070-01DUP	1	Water	NA	NA	19.2	5.24	12/11/2020	15:11
8	CCV2	1	Water	NA	NA	21.3	12.02	12/11/2020	15:12

2B 112334

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L5070 WorkList ID : 144874 Department : Wet-Chemistry Date : 12-11-2020 14:49:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/08/2020	Water	L5070-01 -A	pH	Cool 4 deg C	STER15	L51	GILROY-OMEGA-CELL	12/08/2020	9040C

Date/Time 12-11-2020 14:50
Received by: J. Avel-Ortiz
Relinquished by: J. Avel-Ortiz

Date/Time 12-11-2020 16:40
Received by: J. Avel-Ortiz
Relinquished by: J. Avel-Ortiz