


LB112338

ISM002.4 INSTRUMENT ID : CV1

Linear

 μ Abs.:

47155



A= 0.0000e+000

B= 2.1336e-004 *slope*C= -3.3060e-003 *y-intercept*

Rho= 0.9999181

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/e))</i>
0.0	0.000	0.013	0.013	76	0.000	76					-
0.2	0.200	0.205	0.005	975	0.0 %	975					3
2.5	2.500	2.545	0.045	11942	0.0 %	11942					2
5.0	5.000	4.921	-0.079	23081	0.0 %	23081	0				-2
7.5	7.500	7.459	-0.041	34976	0.0 %	34976					-1
10.0	10.000	10.058	0.058	47155	0.0 %	47155					1

LB112338

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	76	-		0 ISM01.3	PPB	12/11/2020 16:21	Std
	0.2 S0.2	975	-		0.2 ISM01.3	PPB	12/11/2020 16:23	Std
	2.5 S2.5	11942	-		2.5 ISM01.3	PPB	12/11/2020 16:25	Std
	5 S5.0	23081	-		5 ISM01.3	PPB	12/11/2020 16:27	Std
	7.5 S7.5	34976	-		7.5 ISM01.3	PPB	12/11/2020 16:29	Std
	10 S10.0	47155	-		10 ISM01.3	PPB	12/11/2020 16:31	Std
ICV	ICV	18118	3.8623 -		ISM01.3	PPB	12/11/2020 16:34	SMPL
ICB	ICB	-27	-0.0091 -		ISM01.3	PPB	12/11/2020 16:36	SMPL
CCV068	CCV068	22724	4.845 -		ISM01.3	PPB	12/11/2020 16:38	SMPL
CCB068	CCB068	-26	-0.0089 -		ISM01.3	PPB	12/11/2020 16:40	SMPL
PB133563BL	PBS563	229	0.0456 -		ISM01.3	PPB	12/11/2020 16:42	SMPL
L5064-01	MC0AD5	2108	0.4465 -		ISM01.3	PPB	12/11/2020 16:44	SMPL
L5064-02	MC0AD5D	2265	0.4799 -		ISM01.3	PPB	12/11/2020 16:46	SMPL
L5064-03	MC0AD5S	14598	3.1113 -		ISM01.3	PPB	12/11/2020 16:53	SMPL
L5064-04	MC0AD6	1736	0.3671 -		ISM01.3	PPB	12/11/2020 16:58	SMPL
L5064-05	MC0AD7	4025	0.8555 -		ISM01.3	PPB	12/11/2020 17:00	SMPL
L5064-06	MC0AD8	2869	0.6088 -		ISM01.3	PPB	12/11/2020 17:03	SMPL
L5064-07	MC0AD9	2023	0.4283 -		ISM01.3	PPB	12/11/2020 17:05	SMPL
L5064-08	MC0AE0	1251	0.2636 -		ISM01.3	PPB	12/11/2020 17:07	SMPL
CCV069	CCV069	23097	4.9246 -		ISM01.3	PPB	12/11/2020 17:09	SMPL
CCB069	CCB069	-21	-0.0078 -		ISM01.3	PPB	12/11/2020 17:11	SMPL