



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112347

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	12/12/2020	09:15
2	CAL2	1	Water	NA	NA	21.3	7.01	12/12/2020	09:16
3	CAL3	1	Water	NA	NA	21.3	10.03	12/12/2020	09:17
4	ICV	1	Water	NA	NA	21.3	7.00	12/12/2020	09:19
5	CCV1	1	Water	NA	NA	21.3	2.01	12/12/2020	09:20
6	L5072-01	1	Solid	20.03	20	21.6	9.23	12/12/2020	09:23
7	L5072-01DUP	1	Solid	20.04	20	21.7	9.24	12/12/2020	09:24
8	L5072-02	1	Solid	20.01	20	21.4	9.98	12/12/2020	09:25
9	L5072-03	1	Solid	20.03	20	21.5	9.04	12/12/2020	09:29
10	L5072-04	1	Solid	20.01	20	21.5	8.30	12/12/2020	09:30
11	CCV2	1	Water	NA	NA	21.3	12.02	12/12/2020	09:32

WORKLIST(Hardcopy Internal Chain)

LB 112347

WorkList Name : PH L5072 WorkList ID : 144892 Department : Wet-Chemistry Date : 12-12-2020 09:05:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/10/2020	Solid	L5072-01	-C pH	Cool 4 deg C	GEIC06	L11	BIN-1-GF-MAPLE	12/10/2020	9045D
12/10/2020	Solid	L5072-02	-C pH	Cool 4 deg C	GEIC06	L11	BIN-3-UU-GALAXY	12/10/2020	9045D
12/10/2020	Solid	L5072-03	-C pH	Cool 4 deg C	GEIC06	L11	BIN-6-TOPSOIL	12/10/2020	9045D
12/10/2020	Solid	L5072-04	-C pH	Cool 4 deg C	GEIC06	L11	BIN-7-COMPOST	12/10/2020	9045D

Date/Time 12.12.2020 9:10
 Received by: JG Galt Perry
 Relinquished by: JG Galt Perry

Date/Time 12.12.2020 10:00
 Received by: JP Custer Perry
 Relinquished by: JP Galt Perry