



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112369

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	12/14/2020	12:10
2	CAL2	1	Water	NA	NA	21.4	7.00	12/14/2020	12:11
3	CAL3	1	Water	NA	NA	21.4	10.03	12/14/2020	12:12
4	ICV	1	Water	NA	NA	21.4	6.99	12/14/2020	12:15
5	CCV1	1	Water	NA	NA	21.4	2.01	12/14/2020	12:16
6	L5086-01	1	Water	NA	NA	21.0	5.68	12/14/2020	12:19
7	L5086-01DUP	1	Water	NA	NA	21.1	5.69	12/14/2020	12:20
8	CCV2	1	Water	NA	NA	21.4	12.02	12/14/2020	12:22

2B 112369

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L5086 WorkList ID : 144942 Department : Wet-Chemistry Date : 12-14-2020 11:52:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/09/2020	Water	L-5086-01	pH	Cool 4 deg C	STER14	L61	JS-SYSTEM	12/09/2020	9040C

Date/Time 12-14-2020 12:10:00
Received by: Jo [Signature]
Relinquished by: CP [Signature]

Date/Time 12-13-2020 13:00
Received by: CP [Signature]
Relinquished by: Jo [Signature]