

LB 112374


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

54678



A= 0.0000e+000

B= 1.8408e-004

C= -7.9881e-003

Rho= 0.9998068

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	125	0.000	125				
0.2	0.200	0.190	-0.010	1077	0.0 %	1077				
2.5	2.500	2.479	-0.021	13511	0.0 %	13511				
5.0	5.000	5.095	0.095	27721	0.0 %	27721				
7.5	7.500	7.364	-0.136	40047	0.0 %	40047				
10.0	10.000	10.057	0.057	54678	0.0 %	54678				

LB112374 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Dec 2020 16:00:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	125	-	0.0000	7471B	PPB	14 Dec 2020 16:05:51	S	Std
0.2 - 1	50.2	1077	-	0.2000	7471B	PPB	14 Dec 2020 16:08:09	S	Std
2.5 - 1	52.5	13511	-	2.5000	7471B	PPB	14 Dec 2020 16:10:26	S	Std
5.0 - 1	55	27721	-	5.0000	7471B	PPB	14 Dec 2020 16:12:43	S	Std
7.5 - 1	59.5	40047	-	7.5000	7471B	PPB	14 Dec 2020 16:15:00	S	Std
10.0 - 1	510	54678	-	10.0000	7471B	PPB	14 Dec 2020 16:19:49	S	Std
ICV40 - 1	ICV40	20117	3.6951	-	7471B	PPB	14 Dec 2020 16:22:33	U	SMPL
ICB40 - 1	ICB40	15	-0.0052	-	7471B	PPB	14 Dec 2020 16:24:48	U	SMPL
CCV75 - 1	CCV75	27691	5.0893	-	7471B	PPB	14 Dec 2020 16:27:03	U	SMPL
CCB75 - 1	CCB75	7	-0.0067	-	7471B	PPB	14 Dec 2020 16:29:18	U	SMPL
CRA - 1	CRA	1051	0.1855	-	7471B	PPB	14 Dec 2020 16:31:33	U	SMPL
HighStd - 1	HighStd	55150	10.1439	-	7471B	PPB	14 Dec 2020 16:33:48	U	SMPL
ChkStd - 1	ChkStd	38229	7.0291	-	7471B	PPB	14 Dec 2020 16:36:04	U	SMPL
PB133594BL - 1	PBS	-72	-0.0212	-	7471B	PPB	14 Dec 2020 16:38:22	U	SMPL
PB133594BS - 1	LCSS	19134	3.5142	-	7471B	PPB	14 Dec 2020 16:40:39	U	SMPL
L5042-02 - 1	KENS-B6-121020-0-6	1960	0.3528	-	7471B	PPB	14 Dec 2020 16:42:56	U	SMPL
L5042-02DUP - 1	KENS-B6-121020-0-6DUP	1703	0.3055	-	7471B	PPB	14 Dec 2020 16:45:12	U	SMPL
L5042-04 - 1	L5042-02MS	20103	3.6925	-	7471B	PPB	14 Dec 2020 16:47:29	U	SMPL
L5042-06 - 1	L5042-02MSD	21218	3.8978	-	7471B	PPB	14 Dec 2020 16:49:46	U	SMPL
L5042-08 - 1	KENS-B6-121020-13-15	244	0.0369	-	7471B	PPB	14 Dec 2020 16:52:03	U	SMPL
CCV76 - 1	CCV76	27929	5.1331	-	7471B	PPB	14 Dec 2020 16:54:18	U	SMPL
CCB76 - 1	CCB76	33	-0.0019	-	7471B	PPB	14 Dec 2020 16:56:34	U	SMPL
L5043-02 - 1	KENS-SS1-121020-0-7	591	0.1008	-	7471B	PPB	14 Dec 2020 16:58:49	U	SMPL
L5043-06 - 1	KENS-SS2-121020-0-7	656	0.1128	-	7471B	PPB	14 Dec 2020 17:01:04	U	SMPL
L5043-07 - 1	L5043-06MS	20448	3.7560	-	7471B	PPB	14 Dec 2020 17:03:19	U	SMPL
L5043-08 - 1	L5043-06MSD	20444	3.7553	-	7471B	PPB	14 Dec 2020 17:05:35	U	SMPL
L5043-10 - 1	KENS-SS-121020-BD	802	0.1396	-	7471B	PPB	14 Dec 2020 17:07:51	U	SMPL
L5045-02 - 1	RBR-10	2143	0.3865	-	7471B	PPB	14 Dec 2020 17:10:08	U	SMPL
L5045-04 - 1	RBR-78	3048	0.5531	-	7471B	PPB	14 Dec 2020 17:12:25	U	SMPL
L5045-06 - 1	RBR-24	1272	0.2262	-	7471B	PPB	14 Dec 2020 17:14:41	U	SMPL
L5047-01 - 1	OK-03-121120	942	0.1654	-	7471B	PPB	14 Dec 2020 17:16:58	U	SMPL
L5048-01 - 1	OR-02-121120	1590	0.2847	-	7471B	PPB	14 Dec 2020 17:19:14	U	SMPL
CCV77 - 1	CCV77	28340	5.2088	-	7471B	PPB	14 Dec 2020 17:21:30	U	SMPL
CCB77 - 1	CCB77	16	-0.0050	-	7471B	PPB	14 Dec 2020 17:23:45	U	SMPL
L5066-01 - 1	JBD-SB-01-0-52	13202	2.4222	-	7471B	PPB	14 Dec 2020 17:26:01	U	SMPL
L5066-02 - 1	JBD-SB-0203-0-15	26426	4.8565	-	7471B	PPB	14 Dec 2020 17:28:17	U	SMPL
L5066-03 - 1	JBD-IB-0809A-0-6	11991	2.1993	-	7471B	PPB	14 Dec 2020 17:30:32	U	SMPL
L5066-04 - 1	JBD-BOLLARDS-0-6	31131	5.7225	-	7471B	PPB	14 Dec 2020 17:32:48	U	SMPL
L5066-05 - 1	L5066-04MS	43101	7.9259	-	7471B	PPB	14 Dec 2020 17:35:04	U	SMPL
L5066-06 - 1	L5066-04MSD	47802	8.7913	-	7471B	PPB	14 Dec 2020 17:37:20	U	SMPL
L5069-02 - 1	KENS-B7-121120-0-4	1134	0.2008	-	7471B	PPB	14 Dec 2020 17:39:37	U	SMPL
L5069-04 - 1	KENS-B7-121120-13-15	200	0.0288	-	7471B	PPB	14 Dec 2020 17:41:56	U	SMPL
L5042-02LX5 - 1		352	0.0568	-	7471B	PPB	14 Dec 2020 17:44:13	U	SMPL
L5042-02A - 1		23166	4.2564	-	7471B	PPB	14 Dec 2020 17:46:28	U	SMPL
CCV78 - 1	CCV78	28204	5.1837	-	7471B	PPB	14 Dec 2020 17:48:44	U	SMPL
CCB78 - 1	CCB78	45	0.0003	-	7471B	PPB	14 Dec 2020 17:50:59	U	SMPL