

LB112375

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Am

Instrument ID : Konelab

12/14/2020 16:49

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.668	0.0	0.085	
ICB1	0.520	0.0	0.001	
CCV1	241.307	0.0	0.214	
CCB1	0.660	0.0	0.001	
L5028-07	8.504	0.0	0.008	
L5028-07DUP	8.620	0.0	0.008	
L5028-07MS	33.188	0.0	0.030	
L5028-07MSD	33.316	0.0	0.030	
L5051-07	8.436	0.0	0.008	
L5076-02	0.223	0.0	0.001	
MIDPB133556	246.644	0.0	0.219	
PB133556BL	-0.100	0.0	0.000	
PB133566BS	99.734	0.0	0.089	
L5028-07A	47.437	0.0	0.043	
CCV2	253.157	0.0	0.225	
CCB2	0.275	0.0	0.001	

N	16
Mean	67.349
SD	94.6417
CV%	140.52

Aquakem v. 7.2AQ1

Results from time period:

Mon Dec 14 15:36:16 2020

Mon Dec 14 16:44:14 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.6339	µg/l	12/14/2020 13:07:37	
5.OPPBCN	A	Total CN	P	4.5521	µg/l	12/14/2020 13:07:38	
10PPBCN	A	Total CN	P	9.6251	µg/l	12/14/2020 13:07:39	
50PPBCN	A	Total CN	P	50.102	µg/l	12/14/2020 13:07:40	
100PPBCN	A	Total CN	P	100.7885	µg/l	12/14/2020 13:07:41	
250PPBCN	A	Total CN	P	251.4442	µg/l	12/14/2020 13:07:42	
500PPBCN	A	Total CN	P	499.122	µg/l	12/14/2020 13:07:43	
ICV1	S	Total CN	P	95.6679	µg/l	12/14/2020 16:13:08	
ICB1	S	Total CN	P	0.5198	µg/l	12/14/2020 16:13:09	
CCV1	S	Total CN	P	241.3068	µg/l	12/14/2020 16:13:10	
CCB1	S	Total CN	P	0.6596	µg/l	12/14/2020 16:13:11	
L5028-07	S	Total CN	P	8.5041	µg/l	12/14/2020 16:19:26	
L5028-07DUP	S	Total CN	P	8.6204	µg/l	12/14/2020 16:19:27	
L5028-07MS	S	Total CN	P	33.1883	µg/l	12/14/2020 16:19:28	
L5028-07MSD	S	Total CN	P	33.3158	µg/l	12/14/2020 16:19:29	
L5051-07	S	Total CN	P	8.4357	µg/l	12/14/2020 16:19:30	
L5076-02	S	Total CN	P	0.223	µg/l	12/14/2020 16:19:31	
MIDPB133556	S	Total CN	P	246.6444	µg/l	12/14/2020 16:24:34	
PB133556BL	S	Total CN	P	-0.1	µg/l	12/14/2020 16:44:10	
PB133566BS	S	Total CN	P	99.7335	µg/l	12/14/2020 16:44:11	
L5028-07A	S	Total CN	P	47.4373	µg/l	12/14/2020 16:44:12	
CCV2	S	Total CN	P	253.1574	µg/l	12/14/2020 16:44:13	
CCB2	S	Total CN	P	0.2746	µg/l	12/14/2020 16:44:14	

LB112375

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

12/14/2020 13:11


Test Total CN

Accepted 12/14/2020 13:11

Factor 1129
Bias 0.001

Coeff. of det. 0.999979

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	-0.000	-0.6339	0.0000	
2	5.0PPBCN	0.005	4.5521	5.0000	
3	10PPBCN	0.009	9.6251	10.0000	
4	50PPBCN	0.045	50.1020	50.0000	
5	100PPBCN	0.090	100.7885	100.0000	
6	250PPBCN	0.223	251.4442	250.0000	
7	500PPBCN	0.443	499.1220	500.0000	