



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112379

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPONT

Barometric Scale ID: AZ-55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	12/14/2020	10:50
2	L5039-01		1	9	100.00	>212.0	>212.0	12/14/2020	11:25
3	L5039-01DUP		1	9	100.00	>212.0	>212.0	12/14/2020	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112379

WorkList Name : fp121320

WorkList ID : 144924

Department : Wet-Chemistry

Date : 12-13-2020 13:14:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/17/2020	Water	L5039-01	Flash Point		PARS15	L51	JBD-WC-L-SB-02	12/10/2020	1010A

Date/Time 12/14/20 10:20
 Received by: KS (Wetchem)
 Relinquished by: SA (Wetchem)

Date/Time 12/14/20 11:50
 Received by: JO (Wetchem)
 Relinquished by: KS (Wetchem)