


LB112385

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.6688e-004

C= -2.3385e-003

Rho= 0.9997834

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	41	0.000	41				
0.2	0.200	0.199	-0.001	754	0.0 %	754				
2.5	2.500	2.395	-0.105	8983	0.0 %	8983				
5.0	5.000	5.121	0.121	19199	0.0 %	19199				
7.5	7.500	7.556	0.056	28321	0.0 %	28321				
10.0	10.000	9.920	-0.080	37179	0.0 %	37179				

LB112385 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Dec 2020 11:18:18

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7470A	PPB	15 Dec 2020 11:29:05	S	Std
0.2 - 1	50.2	754	-	0.2000	7470A	PPB	15 Dec 2020 11:31:22	S	Std
2.5 - 1	52.5	8983	-	2.5000	7470A	PPB	15 Dec 2020 11:33:40	S	Std
5.0 - 1	55	19199	-	5.0000	7470A	PPB	15 Dec 2020 11:35:57	S	Std
7.5 - 1	57.5	28321	-	7.5000	7470A	PPB	15 Dec 2020 11:38:13	S	Std
10.0 - 1	510	37179	-	10.0000	7470A	PPB	15 Dec 2020 11:40:29	S	Std
ICV41 - 1	ICV41	13679	3.6483	-	7470A	PPB	15 Dec 2020 11:43:12	U	SMPL
ICB41 - 1	ICB41	-49	-0.0154	-	7470A	PPB	15 Dec 2020 11:45:26	U	SMPL
CCV79 - 1	CCV79	18742	4.9995	-	7470A	PPB	15 Dec 2020 11:47:42	U	SMPL
CCB79 - 1	CCB79	-46	-0.0146	-	7470A	PPB	15 Dec 2020 11:49:57	U	SMPL
CRA - 1	CRA	742	0.1957	-	7470A	PPB	15 Dec 2020 11:52:12	U	SMPL
HighStd - 1	HighStd	37194	9.9240	-	7470A	PPB	15 Dec 2020 11:54:27	U	SMPL
ChkStd - 1	ChkStd	27176	7.2504	-	7470A	PPB	15 Dec 2020 11:56:42	U	SMPL
PB133611BL - 1	PBW	-78	-0.0232	-	7470A	PPB	15 Dec 2020 11:58:58	U	SMPL
PB133611BS - 1	LCSW	13813	3.6841	-	7470A	PPB	15 Dec 2020 12:01:15	U	SMPL
L5022-01 - 1	WCS-OH-E6	136	0.0340	-	7470A	PPB	15 Dec 2020 12:03:33	U	SMPL
L5059-04 - 1	CALDRUM1-20201209	17	0.0022	-	7470A	PPB	15 Dec 2020 12:05:49	U	SMPL
L5065-01 - 1	JBD-WC-S-SB-01-0-52	35	0.0070	-	7470A	PPB	15 Dec 2020 12:08:06	U	SMPL
L5065-02 - 1	JBD-WC-S-SB-0203-0-15	42	0.0089	-	7470A	PPB	15 Dec 2020 12:10:23	U	SMPL
L5065-03 - 1	JBD-WC-S-IB-0809A-0-6	37	0.0075	-	7470A	PPB	15 Dec 2020 12:12:38	U	SMPL
CCV80 - 1	CCV80	18668	4.9798	-	7470A	PPB	15 Dec 2020 12:14:53	U	SMPL
CCB80 - 1	CCB80	-34	-0.0114	-	7470A	PPB	15 Dec 2020 12:17:09	U	SMPL
L5065-04 - 1	JBD-WC-S-BOLLARDS-0-6	46	0.0099	-	7470A	PPB	15 Dec 2020 12:19:24	U	SMPL
L5065-04DUP - 1	JBD-WC-S-BOLLARDS-0-6DUP	25	0.0043	-	7470A	PPB	15 Dec 2020 12:21:40	U	SMPL
L5065-04MS - 1	JBD-WC-S-BOLLARDS-0-6MS	13978	3.7281	-	7470A	PPB	15 Dec 2020 12:23:55	U	SMPL
L5065-04MSD - 1	JBD-WC-S-BOLLARDS-0-6MSD	14191	3.7850	-	7470A	PPB	15 Dec 2020 12:26:11	U	SMPL
PB133612BL - 1	PBW	-27	-0.0095	-	7470A	PPB	15 Dec 2020 12:28:27	U	SMPL
PB133612BS - 1	LCSW	13065	3.4844	-	7470A	PPB	15 Dec 2020 12:30:44	U	SMPL
L5059-03 - 1	CALTANK2-20201209	-15	-0.0063	-	7470A	PPB	15 Dec 2020 12:33:00	U	SMPL
L5059-03DUP - 1	CALTANK2-20201209DUP	31	0.0059	-	7470A	PPB	15 Dec 2020 12:35:17	U	SMPL
L5059-03MS - 1	CALTANK2-20201209MS	13384	3.5696	-	7470A	PPB	15 Dec 2020 12:37:34	U	SMPL
L5059-03MSD - 1	CALTANK2-20201209MSD	14020	3.7393	-	7470A	PPB	15 Dec 2020 12:39:49	U	SMPL
CCV81 - 1	CCV81	18489	4.9320	-	7470A	PPB	15 Dec 2020 12:42:05	U	SMPL
CCB81 - 1	CCB81	-44	-0.0141	-	7470A	PPB	15 Dec 2020 12:44:20	U	SMPL
PB133611TB - 1	PB133564TB	23	0.0038	-	7470A	PPB	15 Dec 2020 12:46:36	U	SMPL
L5065-04LX5 - 1		25	0.0043	-	7470A	PPB	15 Dec 2020 12:48:52	U	SMPL
L5065-04A - 1		15928	4.2485	-	7470A	PPB	15 Dec 2020 12:51:07	U	SMPL
L5059-03LX5 - 1		-24	-0.0087	-	7470A	PPB	15 Dec 2020 12:53:23	U	SMPL
L5059-03A - 1		14968	3.9923	-	7470A	PPB	15 Dec 2020 13:05:18	U	SMPL
CCV82 - 1	CCV82	19181	5.1167	-	7470A	PPB	15 Dec 2020 13:07:34	U	SMPL
CCB82 - 1	CCB82	-48	-0.0151	-	7470A	PPB	15 Dec 2020 13:09:50	U	SMPL