



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112390

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	12/15/2020	15:10
2	CAL2	1	Water	NA	NA	21.3	7.00	12/15/2020	15:11
3	CAL3	1	Water	NA	NA	21.3	10.03	12/15/2020	15:12
4	ICV	1	Water	NA	NA	21.3	6.99	12/15/2020	15:15
5	CCV1	1	Water	NA	NA	21.3	2.01	12/15/2020	15:16
6	L5115-02	1	Water	NA	NA	13.2	6.59	12/15/2020	15:18
7	L5115-02DUP	1	Water	NA	NA	13.3	6.60	12/15/2020	15:19
8	CCV2	1	Water	NA	NA	21.4	12.02	12/15/2020	15:20

183 112390

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L5115 W. WorkList ID : 144978 Department : Wet-Chemistry Date : 12-15-2020 14:55:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2020	Water	L5115-02	Corrosivity	Cool 4 deg C	SEAC01	L61	GROUNDWATER	12/15/2020	9040C

Date/Time 12-15-2020 15:00
Received by: Jo Wet Chem
Relinquished by: KS (watcher)

Date/Time 12-15-2020 16:00
Received by: KS (watcher)
Relinquished by: JP Wet Chem