



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112393

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	12/15/2020	15:10
2	CAL2	1	Water	NA	NA	21.3	7.01	12/15/2020	15:11
3	CAL3	1	Water	NA	NA	21.3	10.03	12/15/2020	15:12
4	ICV	1	Water	NA	NA	21.3	7.00	12/15/2020	15:15
5	CCV1	1	Water	NA	NA	21.3	2.01	12/15/2020	15:16
6	L5113-03	1	Solid	20.03	20	21.2	6.65	12/15/2020	15:20
7	L5113-03DUP	1	Solid	20.04	20	21.4	6.66	12/15/2020	15:21
8	L5115-01	1	Solid	20.01	20	20.7	5.70	12/15/2020	15:23
9	CCV2	1	Water	NA	NA	21.4	12.02	12/15/2020	15:25

WORKLIST(Hardcopy Internal Chain)

✓ 112393

WorkList Name : CORROSIVITY L5115 S WorkList ID : 144977 Department : Wet-Chemistry Date : 12-15-2020 14:54:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2020	Solid	L5113-03	Corrosivity	Cool 4 deg C	PSEG03	L53	SS1	12/15/2020	9045D
12/15/2020	Solid	L5115-01	Corrosivity	Cool 4 deg C	SEAC01	L61	SOIL-CUTTINGS	12/15/2020	9045D

Date/Time 12-15-2020 15:00
 Received by: JS (Wetchem)
 Relinquished by: JS (Wetchem)

Date/Time 12-15-2020 18:00
 Received by: JS (Wetchem)
 Relinquished by: JS (Wetchem)