

LB112396


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

3728E



A= 0.0000e+000

B= 2.6852e-004

C= 9.8587e-003

Rho= 0.9999807

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	48	0.000	48					-
0.2	0.200	0.209	0.009	740	0.0 %	740					5
2.5	2.500	2.487	-0.013	9225	0.0 %	9225					-1
5.0	5.000	4.957	-0.043	18422	0.0 %	18422					-1
7.5	7.500	7.503	0.003	27904	0.0 %	27904					0
10.0	10.000	10.022	0.022	37288	0.0 %	37288					0

LB112396 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 16 Dec 2020 09:23:03					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	48	-	0.0000	7470A	PPB	16 Dec 2020 09:23:22	S	Std
0.2 - 1	S0.2	740	-	0.2000	7470A	PPB	16 Dec 2020 09:25:40	S	Std
2.5 - 1	S2.5	9225	-	2.5000	7470A	PPB	16 Dec 2020 09:27:57	S	Std
5.0 - 1	S5	18422	-	5.0000	7470A	PPB	16 Dec 2020 09:30:14	S	Std
7.5 - 1	S7.5	27904	-	7.5000	7470A	PPB	16 Dec 2020 09:32:30	S	Std
10.0 - 1	S10	37288	-	10.0000	7470A	PPB	16 Dec 2020 09:34:46	S	Std
ICV43 - 1	ICV43	13613	3.6652	-	7470A	PPB	16 Dec 2020 09:38:05	U	SMPL
ICB43 - 1	ICB43	-9	0.0074	-	7470A	PPB	16 Dec 2020 09:40:20	U	SMPL
CCV87 - 1	CCV87	18632	5.0129	-	7470A	PPB	16 Dec 2020 09:42:35	U	SMPL
CCB87 - 1	CCB87	-28	0.0023	-	7470A	PPB	16 Dec 2020 09:44:50	U	SMPL
CRA - 1	CRA	703	0.1986	-	7470A	PPB	16 Dec 2020 09:47:05	U	SMPL
HighStd - 1	HighStd	37545	10.0915	-	7470A	PPB	16 Dec 2020 09:49:20	U	SMPL
ChkStd - 1	ChkStd	26726	7.1864	-	7470A	PPB	16 Dec 2020 09:51:36	U	SMPL
PB133627BL - 1	PBW	-72	-0.0095	-	7470A	PPB	16 Dec 2020 09:53:52	U	SMPL
PB133627BS - 1	LCSW	14615	3.9343	-	7470A	PPB	16 Dec 2020 09:56:08	U	SMPL
L5088-02 - 1	TR1-A-D-COMP	-16	0.0056	-	7470A	PPB	16 Dec 2020 09:58:24	U	SMPL
L5088-13 - 1	TR3-A-D-COMP	40	0.0206	-	7470A	PPB	16 Dec 2020 10:00:41	U	SMPL
L5088-13DUP - 1	TR3-A-D-COMPDUP	45	0.0219	-	7470A	PPB	16 Dec 2020 10:02:57	U	SMPL
L5088-13MS - 1	TR3-A-D-COMPMS	13821	3.7211	-	7470A	PPB	16 Dec 2020 10:05:14	U	SMPL
L5088-13MSD - 1	TR3-A-D-COMPMSD	13150	3.5409	-	7470A	PPB	16 Dec 2020 10:07:30	U	SMPL
CCV88 - 1	CCV88	18542	4.9888	-	7470A	PPB	16 Dec 2020 10:09:45	U	SMPL
CCB88 - 1	CCB88	-33	0.0010	-	7470A	PPB	16 Dec 2020 10:12:01	U	SMPL
L5090-02 - 1	DARK-GREEN-CHIPS	40	0.0206	-	7470A	PPB	16 Dec 2020 10:14:16	U	SMPL
L5090-04 - 1	LIGHT-GREEN-CHIPS	71	0.0289	-	7470A	PPB	16 Dec 2020 10:16:32	U	SMPL
L5090-06 - 1	GRAY-CHIPS	56	0.0249	-	7470A	PPB	16 Dec 2020 10:18:48	U	SMPL
PB133627TB - 1	PB133620TB	51	0.0236	-	7470A	PPB	16 Dec 2020 10:21:03	U	SMPL
L5088-13LX5 - 1		181	0.0585	-	7470A	PPB	16 Dec 2020 10:23:19	U	SMPL
L5088-13A - 1		15009	4.0401	-	7470A	PPB	16 Dec 2020 10:25:35	U	SMPL
CCV89 - 1	CCV89	18443	4.9622	-	7470A	PPB	16 Dec 2020 10:27:51	U	SMPL
CCB89 - 1	CCB89	-36	0.0002	-	7470A	PPB	16 Dec 2020 10:30:08	U	SMPL