



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112399

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.00	80.6	80.3	12/16/2020	11:05
2	L5115-02		1	8	100.00	>212.0	>212.0	12/16/2020	11:35
3	L5115-02DUP		1	8	100.00	>212.0	>212.0	12/16/2020	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112399

WorkList Name : fp121620 WorkList ID : 145002 Department : Wet-Chemistry Date : 12-16-2020 08:54:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/28/2020	Water	L5115-02	Flash Point	Cool 4 deg C	SEAC01	L61	GROUNDWATER	12/15/2020	1010A

Date/Time 12/16/20 10:40
 Received by: KS (wetcher)
 Relinquished by: JQ (wetcher)

Date/Time 12/16/20 12:15
 Received by: JQ (wetcher)
 Relinquished by: KS (wetcher)