

LB112400


7470A

INSTRUMENT ID: CYL

Linear

 μ Abs.:

39636



A= 0.0000e+000

B= 2.5265e-004

C= 4.9978e-003

Rho= 0.9999866

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y/D
0.0	0.000	0.025	0.025	80	0.000	80					-
0.2	0.200	0.203	0.003	784	0.0 %	784					2
2.5	2.500	2.483	-0.017	9806	0.0 %	9806					-1
5.0	5.000	4.970	-0.030	19653	0.0 %	19653					-1
7.5	7.500	7.500	-0.000	29664	0.0 %	29664					0
10.0	10.000	10.019	0.019	39636	0.0 %	39636					0

LB112400 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Dec 2020 12:25:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	80	-	0.0000	7470A	PPB	16 Dec 2020 12:26:24	S	Std
0.2 - 1	50.2	784	-	0.2000	7470A	PPB	16 Dec 2020 12:28:41	S	Std
2.5 - 1	52.5	9806	-	2.5000	7470A	PPB	16 Dec 2020 12:30:58	S	Std
5.0 - 1	55	19653	-	5.0000	7470A	PPB	16 Dec 2020 12:33:15	S	Std
7.5 - 1	57.5	29664	-	7.5000	7470A	PPB	16 Dec 2020 12:35:32	S	Std
10.0 - 1	50	39636	-	10.0000	7470A	PPB	16 Dec 2020 12:40:19	S	Std
ICV44 - 1	ICV44	14387	3.6399	-	7470A	PPB	16 Dec 2020 12:43:01	U	SMPL
ICB44 - 1	ICB44	-17	0.0007	-	7470A	PPB	16 Dec 2020 12:45:15	U	SMPL
CCV90 - 1	CCV90	19293	4.8794	-	7470A	PPB	16 Dec 2020 12:47:30	U	SMPL
CCB90 - 1	CCB90	-25	-0.0013	-	7470A	PPB	16 Dec 2020 12:49:45	U	SMPL
CRA - 1	CRA	735	0.1907	-	7470A	PPB	16 Dec 2020 12:52:01	U	SMPL
HighStd - 1	HighStd	38818	9.8125	-	7470A	PPB	16 Dec 2020 12:54:16	U	SMPL
ChkStd - 1	ChkStd	27321	6.9077	-	7470A	PPB	16 Dec 2020 12:56:32	U	SMPL
PB133646BL - 1	PBW	-51	-0.0079	-	7470A	PPB	16 Dec 2020 12:58:48	U	SMPL
L5113-03 - 1	SS1	-7	0.0032	-	7470A	PPB	16 Dec 2020 13:03:20	U	SMPL
L5113-03DUP - 1	SS1DUP	39	0.0149	-	7470A	PPB	16 Dec 2020 13:05:37	U	SMPL
L5113-03MS - 1	SS1MS	14668	3.7109	-	7470A	PPB	16 Dec 2020 13:10:26	U	SMPL
L5113-03MSD - 1	SS1MSD	14612	3.6968	-	7470A	PPB	16 Dec 2020 13:12:43	U	SMPL
L5115-01 - 1	SOIL-CUTTINGS	-9	0.0027	-	7470A	PPB	16 Dec 2020 13:14:59	U	SMPL
CCV91 - 1	CCV91	18952	4.7933	-	7470A	PPB	16 Dec 2020 13:17:14	U	SMPL
CCB91 - 1	CCB91	-45	-0.0064	-	7470A	PPB	16 Dec 2020 13:19:29	U	SMPL
PB133647BL - 1	PBW	34	0.0136	-	7470A	PPB	16 Dec 2020 13:21:45	U	SMPL
PB133647BS - 1	LCSW	14926	3.7761	-	7470A	PPB	16 Dec 2020 13:24:00	U	SMPL
L5111-01 - 1	SK-12-1-SPA	173	0.0487	-	7470A	PPB	16 Dec 2020 13:26:15	U	SMPL
L5111-01DUP - 1	SK-12-1-SPADUP	159	0.0452	-	7470A	PPB	16 Dec 2020 13:28:31	U	SMPL
L5111-01MS - 1	SK-12-1-SPAMS	14283	3.6136	-	7470A	PPB	16 Dec 2020 13:30:47	U	SMPL
L5111-01MSD - 1	SK-12-1-SPAMSD	13981	3.5373	-	7470A	PPB	16 Dec 2020 13:33:03	U	SMPL
L5111-02 - 1	SK-12-1-SPC	154	0.0439	-	7470A	PPB	16 Dec 2020 13:35:19	U	SMPL
L5111-03 - 1	SK-12-1-SPD	106	0.0318	-	7470A	PPB	16 Dec 2020 13:37:36	U	SMPL
PB133648BL - 1	PBW	22	0.0106	-	7470A	PPB	16 Dec 2020 13:39:53	U	SMPL
PB133648BS - 1	LCSW	13696	3.4653	-	7470A	PPB	16 Dec 2020 13:42:09	U	SMPL
CCV92 - 1	CCV92	19059	4.8203	-	7470A	PPB	16 Dec 2020 13:44:25	U	SMPL
CCB92 - 1	CCB92	-37	-0.0044	-	7470A	PPB	16 Dec 2020 13:46:40	U	SMPL
L5036-12 - 1	WP-1	14726	3.7256	-	7470A	PPB	16 Dec 2020 13:51:38	U	SMPL
L5115-02 - 1	GROUNDWATER	-134	-0.0289	-	7470A	PPB	16 Dec 2020 13:53:53	U	SMPL
L5115-02DUP - 1	GROUNDWATERDUP	23	0.0108	-	7470A	PPB	16 Dec 2020 13:56:11	U	SMPL
L5115-02MS - 1	GROUNDWATERMS	13454	3.4042	-	7470A	PPB	16 Dec 2020 13:58:27	U	SMPL
L5115-02MSD - 1	GROUNDWATERMSD	13705	3.4676	-	7470A	PPB	16 Dec 2020 14:00:43	U	SMPL
PB133646BS - 1	LCSW	14826	3.7508	-	7470A	PPB	16 Dec 2020 14:05:42	U	SMPL
PB133646TB - 1	PB133631TB	0	0.0050	-	7470A	PPB	16 Dec 2020 14:07:59	U	SMPL
PB133647TB - 1	PB133633TB	26	0.0116	-	7470A	PPB	16 Dec 2020 14:10:15	U	SMPL
L5113-03LX5 - 1		45	0.0164	-	7470A	PPB	16 Dec 2020 14:12:32	U	SMPL
L5113-03A - 1		16380	4.1434	-	7470A	PPB	16 Dec 2020 14:14:48	U	SMPL
CCV93 - 1	CCV93	18503	4.6798	-	7470A	PPB	16 Dec 2020 14:17:03	U	SMPL
CCB93 - 1	CCB93	-38	-0.0046	-	7470A	PPB	16 Dec 2020 14:19:19	U	SMPL
L5111-01LX5 - 1		38	0.0146	-	7470A	PPB	16 Dec 2020 14:21:34	U	SMPL
L5111-01A - 1		16551	4.1867	-	7470A	PPB	16 Dec 2020 14:23:50	U	SMPL
L5115-02LX5 - 1		6	0.0065	-	7470A	PPB	16 Dec 2020 14:26:05	U	SMPL
L5115-02A - 1		16421	4.1538	-	7470A	PPB	16 Dec 2020 14:28:21	U	SMPL
CCV94 - 1	CCV94	18309	4.6308	-	7470A	PPB	16 Dec 2020 14:30:37	U	SMPL
CCB94 - 1	CCB94	-51	-0.0079	-	7470A	PPB	16 Dec 2020 14:32:53	U	SMPL