



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 12/17/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 16:42  
In Date: 12/16/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:11  
Out Date: 12/17/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB112402

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments    |
|----------|-----------------|--------|----------------|-------------------------|---------------------------|---------|-------------|
| L5110-01 | A54L2           | 1      | 1.12           | 9.61                    | 8.27                      | 84.2    |             |
| L5110-02 | A54L3           | 2      | 1.15           | 9.66                    | 8.35                      | 84.6    |             |
| L5110-03 | A54L4           | 3      | 1.16           | 9.7                     | 9.15                      | 93.6    |             |
| L5110-04 | A54L5           | 4      | 1.18           | 9.62                    | 8.7                       | 89.1    |             |
| L5110-05 | A54L6           | 5      | 1.14           | 9.66                    | 9.18                      | 94.4    |             |
| L5110-06 | A54L7           | 6      | 1.14           | 9.97                    | 8.87                      | 87.5    |             |
| L5110-07 | A54L8           | 7      | 1.15           | 9.96                    | 8.99                      | 89.0    |             |
| L5110-08 | A54L9           | 8      | 1.2            | 9.58                    | 9.03                      | 93.4    |             |
| L5110-09 | A54M0           | 9      | 1.15           | 9.97                    | 9.12                      | 90.4    |             |
| L5110-10 | A54M1           | 10     | 1.13           | 9.65                    | 8.85                      | 90.6    |             |
| L5110-11 | A54M2           | 11     | 1.13           | 9.61                    | 9.04                      | 93.3    |             |
| L5110-12 | A54M2MS         | 12     | 1.13           | 9.61                    | 9.04                      | 93.3    |             |
| L5110-13 | A54M2MSD        | 13     | 1.13           | 9.61                    | 9.04                      | 93.3    |             |
| L5110-14 | A54M3           | 14     | 1.16           | 9.96                    | 9.1                       | 90.2    |             |
| L5110-15 | A54M4           | 15     | 1.13           | 9.81                    | 9.38                      | 95.0    |             |
| L5110-16 | A54M5           | 16     | 1.13           | 9.7                     | 9.17                      | 93.8    |             |
| L5110-17 | A54M6           | 17     | 1.17           | 9.92                    | 9.22                      | 92.0    |             |
| L5110-18 | A54M7           | 18     | 1.16           | 9.78                    | 8.71                      | 87.6    |             |
| L5110-19 | A54M8           | 19     | 1.13           | 9.53                    | 8.79                      | 91.2    |             |
| L5110-20 | A54M9           | 20     | 1.16           | 9.79                    | 7.91                      | 78.2    |             |
| L5110-21 | A54N0           | 21     | 1.16           | 9.77                    | 8.53                      | 85.6    |             |
| L5110-22 | A54N1           | 22     | 1.12           | 9.92                    | 8.26                      | 81.1    |             |
| L5110-23 | A54P4           | 23     | 1.00           | 2.00                    | 2.00                      | 100.0   | P.E. SAMPLE |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$