


LB 112411

7471B

INSTRUMENT ID: CV2

μ Abs.:

54144



Linear

A= 0.0000e+000

B= 1.8360e-004

C= 7.3161e-003

Rho= 0.9999337

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	36	0.000	36				
0.2	0.200	0.194	-0.006	1015	0.0 %	1015				
2.5	2.500	2.447	-0.053	13287	0.0 %	13287				
5.0	5.000	5.036	0.036	27387	0.0 %	27387				
7.5	7.500	7.562	0.062	41146	0.0 %	41146				
10.0	10.000	9.948	-0.052	54144	0.0 %	54144				

LB112411 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Dec 2020 12:25:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	36	-	0.00007471B	PPB		17 Dec 2020 12:25:52	S	Std
0.2 - 1	50.2	1015	-	0.20007471B	PPB		17 Dec 2020 12:28:09	S	Std
2.5 - 1	52.5	13287	-	2.50007471B	PPB		17 Dec 2020 12:30:26	S	Std
5.0 - 1	55	27387	-	5.00007471B	PPB		17 Dec 2020 12:32:43	S	Std
7.5 - 1	57.5	41146	-	7.50007471B	PPB		17 Dec 2020 12:35:00	S	Std
10.0 - 1	510	54144	-	10.00007471B	PPB		17 Dec 2020 12:39:49	S	Std
ICV45 - 1	ICV45	20780	3.8225	-7471B	PPB		17 Dec 2020 12:42:27	U	SMPL
ICB45 - 1	ICB45	-34	0.0011	-7471B	PPB		17 Dec 2020 12:44:42	U	SMPL
CCV95 - 1	CCV95	27175	4.9967	-7471B	PPB		17 Dec 2020 12:46:57	U	SMPL
CCB95 - 1	CCB95	-54	-0.0026	-7471B	PPB		17 Dec 2020 12:49:13	U	SMPL
CRA - 1	CRA	986	0.1883	-7471B	PPB		17 Dec 2020 12:51:28	U	SMPL
HighStd - 1	HighStd	55117	10.1269	-7471B	PPB		17 Dec 2020 12:53:43	U	SMPL
ChkStd - 1	ChkStd	38031	6.9898	-7471B	PPB		17 Dec 2020 12:55:58	U	SMPL
PB133645BL - 1	PBS	-85	-0.0083	-7471B	PPB		17 Dec 2020 12:58:15	U	SMPL
L5036-02 - 1	B-1	1102	0.2096	-7471B	PPB		17 Dec 2020 13:02:48	U	SMPL
L5036-02DUP - 1	B-1DUP	1011	0.1929	-7471B	PPB		17 Dec 2020 13:05:05	U	SMPL
L5036-02MS - 1	B-1MS	22823	4.1976	-7471B	PPB		17 Dec 2020 13:07:22	U	SMPL
L5036-02MSD - 1	B-1MSD	22743	4.1830	-7471B	PPB		17 Dec 2020 13:09:38	U	SMPL
L5036-03 - 1	B-2	11478	2.1147	-7471B	PPB		17 Dec 2020 13:11:54	U	SMPL
CCV96 - 1	CCV96	27141	4.9904	-7471B	PPB		17 Dec 2020 13:14:10	U	SMPL
CCB96 - 1	CCB96	-67	-0.0050	-7471B	PPB		17 Dec 2020 13:16:25	U	SMPL
L5036-04 - 1	B-3	13225	2.4354	-7471B	PPB		17 Dec 2020 13:18:41	U	SMPL
L5036-05 - 1	B-4	14681	2.7028	-7471B	PPB		17 Dec 2020 13:20:57	U	SMPL
L5036-06 - 1	B-5	2581	0.4812	-7471B	PPB		17 Dec 2020 13:23:13	U	SMPL
L5036-07 - 1	B-6	3563	0.6615	-7471B	PPB		17 Dec 2020 13:25:28	U	SMPL
L5036-08 - 1	B-7	2761	0.5142	-7471B	PPB		17 Dec 2020 13:27:44	U	SMPL
L5036-09 - 1	B-8	3518	0.6532	-7471B	PPB		17 Dec 2020 13:30:01	U	SMPL
L5036-10 - 1	B-9	42940	7.8911	-7471B	PPB		17 Dec 2020 13:32:18	U	SMPL
L5036-11 - 1	B-10	2801	0.5216	-7471B	PPB		17 Dec 2020 13:34:35	U	SMPL
L5113-02 - 1	SS1	4462	0.8265	-7471B	PPB		17 Dec 2020 13:36:52	U	SMPL
L5118-01 - 1	R-LA-37-2-WC	1674	0.3147	-7471B	PPB		17 Dec 2020 13:39:08	U	SMPL
CCV97 - 1	CCV97	27308	5.0211	-7471B	PPB		17 Dec 2020 13:41:24	U	SMPL
CCB97 - 1	CCB97	-38	0.0003	-7471B	PPB		17 Dec 2020 13:43:39	U	SMPL
L5119-01 - 1	B-W	480	0.0954	-7471B	PPB		17 Dec 2020 13:45:55	U	SMPL
L5125-02 - 1	DETENTION-LINE-1	2699	0.5029	-7471B	PPB		17 Dec 2020 13:48:10	U	SMPL
L5125-04 - 1	DETENTION-LINE-2	2661	0.4959	-7471B	PPB		17 Dec 2020 13:50:26	U	SMPL
L5125-06 - 1	DETENTION-LINE-3	3771	0.6997	-7471B	PPB		17 Dec 2020 13:52:42	U	SMPL
PB133645BS - 1	LCSS	21746	3.9999	-7471B	PPB		17 Dec 2020 13:54:57	U	SMPL
L5036-02LX5 - 1		192	0.0426	-7471B	PPB		17 Dec 2020 13:57:14	U	SMPL
L5036-02A - 1		22738	4.1820	-7471B	PPB		17 Dec 2020 13:59:31	U	SMPL
CCV98 - 1	CCV98	27189	4.9992	-7471B	PPB		17 Dec 2020 14:01:48	U	SMPL
CCB98 - 1	CCB98	-32	0.0014	-7471B	PPB		17 Dec 2020 14:04:05	U	SMPL