



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/29/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 12/16/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 12/17/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB112412

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L5074-01	MA54H7	1	1.15	9.95	8.34	81.7	
L5074-02	MA54H9	2	1.18	9.58	8.14	82.9	
L5074-03	MA54J1	3	1.15	9.98	7.95	77.0	
L5074-04	MA54J3	4	1.15	9.98	8.15	79.3	
L5074-05	MA54J5	5	1.12	9.64	7.95	80.2	
L5074-06	MA54J7	6	1.14	9.97	8.19	79.8	
L5074-07	MA54J9	7	1.19	9.58	8.12	82.6	
L5074-08	MA54K1	8	1.1	9.86	8.87	88.7	
L5074-09	MA54K3	9	1.18	9.98	8.67	85.1	
L5074-10	MA54K5	10	1.14	9.96	8.27	80.8	
L5074-11	MA54K5D	11	1.14	9.96	8.27	80.8	
L5074-12	MA54K5S	12	1.14	9.96	8.27	80.8	
L5074-13	MA54K7	13	1.1	9.68	8.97	91.7	
L5074-14	MA54K9	14	1.1	9.88	9.16	91.8	
L5074-15	MA54L1	15	1.16	9.66	8.57	87.2	
L5074-16	MA54L3	16	1.15	9.66	8.35	84.6	
L5074-17	MA54L5	17	1.18	9.62	8.7	89.1	
L5074-18	MA54L7	18	1.14	9.97	8.87	87.5	
L5074-19	MA54L9	19	1.12	9.58	9.03	93.5	
L5074-20	MA54M1	20	1.13	9.65	8.85	90.6	
L5074-21	MA54M3	21	1.16	9.96	9.1	90.2	
L5074-22	MA54M5	22	1.13	9.7	9.17	93.8	
L5074-23	MA54U4	23	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$