



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112414

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	12/17/2020	11:30
2	CAL2	1	Water	NA	NA	21.1	7.00	12/17/2020	11:31
3	CAL3	1	Water	NA	NA	21.1	10.03	12/17/2020	11:32
4	ICV	1	Water	NA	NA	21.2	7.01	12/17/2020	11:34
5	CCV1	1	Water	NA	NA	21.1	2.01	12/17/2020	11:35
6	L5126-04	1	Water	NA	NA	21.4	9.47	12/17/2020	11:40
7	L5126-09	1	Water	NA	NA	20.0	7.18	12/17/2020	11:43
8	L5140-01	1	Water	NA	NA	21.0	8.00	12/17/2020	11:45
9	L5140-01DUP	1	Water	NA	NA	21.1	8.01	12/17/2020	11:46
10	CCV2	1	Water	NA	NA	21.1	12.02	12/17/2020	11:47

112414

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 15140

WorkList ID : 145018

Department : Wet-Chemistry

Date : 12-17-2020 10:52:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2020	Water	L5126-04	pH	Cool 4 deg C	PSEG03	L53	190-ACADEMY-ST-OILY-WAT	12/16/2020	9040C
12/16/2020	Water	L5126-09	pH	Cool 4 deg C	PSEG03	L53	CONTAINMENT-OILY-WATER	12/16/2020	9040C
12/16/2020	Water	L5140-01	pH	Cool 4 deg C	TULL02	P62	GROUNDWATER-SAMPLE-1	12/16/2020	9040C

Date/Time 12-17-2020 11:00
 Received by: JS (wetchem)
 Relinquished by: JS (wetchem)

Date/Time 12-17-2020 14:00
 Received by: JS (wetchem)
 Relinquished by: JS (wetchem)