


LB112417
INSTRUMENT ID: CV1

ISM02.4



Linear

A= 0.0000e+000

B= 2.2582e-004

C= -1.6559e-002

Rho= 0.9998763

Accept=Accepted

*slope
y-intercept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.002	0.002	80	0.000	80					
0.05	0.050	0.048	-0.002	286	0.0 %	286					-4
0.20	0.200	0.189	-0.011	909	0.0 %	909					-6
2.50	2.500	2.488	-0.012	11092	0.0 %	11092					0
5.00	5.000	5.101	0.101	22662	0.0 %	22662					2
7.50	7.500	7.389	-0.111	32795	0.0 %	32795					-1
10.0	10.000	10.033	0.033	44505	0.0 %	44505					0

LB112417
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	80	-		0 ISM02.4	PPB	12/17/2020 14:35	Std
	0.05 S0.05	286	-		0.05 ISM02.4	PPB	12/17/2020 14:38	Std
	0.2 S0.2	909	-		0.2 ISM02.4	PPB	12/17/2020 14:40	Std
	2.5 S2.5	11092	-		2.5 ISM02.4	PPB	12/17/2020 14:42	Std
	5 S5.0	22662	-		5 ISM02.4	PPB	12/17/2020 14:45	Std
	7.5 S7.5	32795	-		7.5 ISM02.4	PPB	12/17/2020 14:47	Std
	10 S10.0	44505	-		10 ISM02.4	PPB	12/17/2020 14:52	Std
ICV	ICV	16197	3.641 -		ISM02.4	PPB	12/17/2020 14:54	SMPL
ICB	ICB	220	0.0331 -		ISM02.4	PPB	12/17/2020 14:57	SMPL
CCV001	CCV001	22623	5.0921 -		ISM02.4	PPB	12/17/2020 14:59	SMPL
CCB001	CCB001	62	-0.0026 -		ISM02.4	PPB	12/17/2020 15:01	SMPL
PB133666BL	PBW666	20	-0.012 -		ISM02.4	PPB	12/17/2020 15:03	SMPL
L5030-01	MEW0T9	88	0.0033 -		ISM02.4	PPB	12/17/2020 15:06	SMPL
L5030-02	MEW0W1	29	-0.01 -		ISM02.4	PPB	12/17/2020 15:08	SMPL
L5030-03	MEW0W5	42	-0.0071 -		ISM02.4	PPB	12/17/2020 15:10	SMPL
L5030-04	MEW0W5D	47	-0.0059 -		ISM02.4	PPB	12/17/2020 15:13	SMPL
L5030-05	MEW0W5S	4438	0.9856 -		ISM02.4	PPB	12/17/2020 15:15	SMPL
CCV002	CCV002	22172	4.9903 -		ISM02.4	PPB	12/17/2020 15:17	SMPL
CCB002	CCB002	91	0.004 -		ISM02.4	PPB	12/17/2020 15:19	SMPL