


LB112442

ISM02.4

INSTRUMENT ID: CV4



Linear

A= 0.0000e+000

B= 2.1301e-004 slope

C= -1.5907e-003

Rho= 0.9999676

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.004	0.004	27	0.000	27					1
0.05	0.050					286					1
0.20	0.200	0.196	-0.004	929	0.0 %	929					-2
2.50	2.500	2.461	-0.039	11561	0.0 %	11561					-2
5.00	5.000	5.057	0.057	23750	0.0 %	23750					1
7.50	7.500	7.502	0.002	35225	0.0 %	35225					0
10.0	10.000	9.979	-0.021	46855	0.0 %	46855					0

LB112442
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	27	-		0 ISM02.4	PPB	12/18/2020 15:50	Std
	0.2 S0.2	929	-		0.2 ISM02.4	PPB	12/18/2020 15:52	Std
	2.5 S2.5	11561	-		2.5 ISM02.4	PPB	12/18/2020 15:57	Std
	5 S5.0	23750	-		5 ISM02.4	PPB	12/18/2020 16:00	Std
	7.5 S7.5	35225	-		7.5 ISM02.4	PPB	12/18/2020 16:02	Std
	10 S10.0	46855	-		10 ISM02.4	PPB	12/18/2020 16:04	Std
ICV	ICV	18198	3.8748 -		ISM02.4	PPB	12/18/2020 16:07	SMPL
ICB	ICB	-45	-0.0112 -		ISM02.4	PPB	12/18/2020 16:09	SMPL
CCV012	CCV012	23820	5.0724 -		ISM02.4	PPB	12/18/2020 16:13	SMPL
CCB012	CCB012	-17	-0.0052 -		ISM02.4	PPB	12/18/2020 16:15	SMPL
PB133694BL	PBS694	20	0.0027 -		ISM02.4	PPB	12/18/2020 16:18	SMPL
L5032-01	MC0AA0	1382	0.2928 -		ISM02.4	PPB	12/18/2020 16:20	SMPL
L5032-02	MC0AA2	1488	0.3154 -		ISM02.4	PPB	12/18/2020 16:22	SMPL
L5032-03	MC0AA2D	1494	0.3167 -		ISM02.4	PPB	12/18/2020 16:24	SMPL
L5032-04	MC0AA2S	13670	2.9103 -		ISM02.4	PPB	12/18/2020 16:27	SMPL
L5032-05	MC0AA3	945	0.1997 -		ISM02.4	PPB	12/18/2020 16:29	SMPL
L5032-06	MC0AA4	1323	0.2802 -		ISM02.4	PPB	12/18/2020 16:31	SMPL
L5032-07	MC0AA5	2890	0.614 -		ISM02.4	PPB	12/18/2020 16:33	SMPL
L5032-08	MC0AA6	1637	0.3471 -		ISM02.4	PPB	12/18/2020 16:36	SMPL
L5032-09	MC0AA7	1728	0.3665 -		ISM02.4	PPB	12/18/2020 16:38	SMPL
L5032-10	MC0AA9	2362	0.5015 -		ISM02.4	PPB	12/18/2020 16:40	SMPL
L5032-11	MC0AB0	2021	0.4289 -		ISM02.4	PPB	12/18/2020 16:42	SMPL
L5032-12	MC0AB2	857	0.181 -		ISM02.4	PPB	12/18/2020 16:45	SMPL
L5032-13	MC0AB3	1273	0.2696 -		ISM02.4	PPB	12/18/2020 16:47	SMPL
L5032-14	MC0AB4	1023	0.2163 -		ISM02.4	PPB	12/18/2020 16:49	SMPL
PB133695BL	PBS695	-10	-0.0037 -		ISM02.4	PPB	12/18/2020 16:51	SMPL
L5135-01	MBFN58	12923	2.7512 -		ISM02.4	PPB	12/18/2020 16:59	SMPL
L5135-02	MBFN58D	12888	2.7437 -		ISM02.4	PPB	12/18/2020 17:01	SMPL
L5135-03	MBFN58S	24372	5.19 -		ISM02.4	PPB	12/18/2020 17:04	SMPL
PB133701BL	PBS701	57	0.0106 -		ISM02.4	PPB	12/18/2020 17:06	SMPL
CCV013	CCV013	24097	5.1314 -		ISM02.4	PPB	12/18/2020 17:08	SMPL
CCB013	CCB013	-33	-0.0086 -		ISM02.4	PPB	12/18/2020 17:10	SMPL
L5152-01	MC0AE3	3595	0.7642 -		ISM02.4	PPB	12/18/2020 17:13	SMPL
L5152-02	MC0AE4	2865	0.6087 -		ISM02.4	PPB	12/18/2020 17:15	SMPL
L5152-03	MC0AE5	4200	0.8931 -		ISM02.4	PPB	12/18/2020 17:17	SMPL
L5152-04	MC0AE5D	4144	0.8811 -		ISM02.4	PPB	12/18/2020 17:19	SMPL
L5152-05	MC0AE5S	15862	3.3772 -		ISM02.4	PPB	12/18/2020 17:22	SMPL
L5152-06	MC0AE6	2194	0.4658 -		ISM02.4	PPB	12/18/2020 17:24	SMPL
L5152-07	MC0AE7	2597	0.5516 -		ISM02.4	PPB	12/18/2020 17:26	SMPL
L5152-08	MC0AE8	3911	0.8315 -		ISM02.4	PPB	12/18/2020 17:29	SMPL
CCV014	CCV014	23874	5.0839 -		ISM02.4	PPB	12/18/2020 17:31	SMPL
CCB014	CCB014	-18	-0.0054 -		ISM02.4	PPB	12/18/2020 17:33	SMPL