



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112448

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/18/2020	16:40
2	CAL2	1	Water	NA	NA	20.1	7.00	12/18/2020	16:41
3	CAL3	1	Water	NA	NA	20.1	10.03	12/18/2020	16:42
4	ICV	1	Water	NA	NA	20.2	7.01	12/18/2020	16:45
5	CCV1	1	Water	NA	NA	20.2	2.01	12/18/2020	16:48
6	L5157-01	1	Water	NA	NA	13.0	5.60	12/18/2020	16:50
7	L5157-01DUP	1	Water	NA	NA	13.1	5.61	12/18/2020	16:51
8	L5157-03	1	Water	NA	NA	13.9	5.62	12/18/2020	17:00
9	CCV2	1	Water	NA	NA	21.2	12.02	12/18/2020	17:02

WORKLIST(Hardcopy Internal Chain)

183 112448

WorkList Name : ph 15157 WorkList ID : 145084 Department : Wet-Chemistry Date : 12-18-2020 16:26:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2020	Water	L5157-01	pH	Cool 4 deg C	PSEG03	L62	32846	12/18/2020	9040C
12/18/2020	Water	L5157-03	pH	Cool 4 deg C	PSEG03	L62	32847	12/18/2020	9040C

Date/Time 12.18.2020 16:30
 Received by: Jo Goff-Carmy
 Relinquished by: F.C. Csm

Date/Time 12.18.2020 18:30
 Received by: F.C. Csm
 Relinquished by: 201604
Ger