



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112470

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPONT

Barometric Scale ID: AZ-55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	12/21/2020	12:20
2	L5157-01		1	9	100.00	>212.0	>212.0	12/21/2020	12:50
3	L5157-01DUP		1	9	100.00	>212.0	>212.0	12/21/2020	13:10
4	L5157-03		1	8	100.00	>212.0	>212.0	12/21/2020	13:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112470

WorkList Name : fp122120 WorkList ID : 145117 Department : Wet-Chemistry Date : 12-21-2020 11:30:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/25/2020	Water	L5157-01	Flash Point	Cool 4 deg C	PSEG03	L62	32846	12/18/2020	1010A
12/25/2020	Water	L5157-03	Flash Point	Cool 4 deg C	PSEG03	L62	32847	12/18/2020	1010A

Date/Time 11/21/20 12:00
 Received by: KS (wekchem)
 Relinquished by: Jor (wekchem)

Date/Time 11/21/20 13:40
 Received by: Jor (wekchem)
 Relinquished by: KS (wekchem)