


LB 112474

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	40	0.000	40				
0.2	0.200	0.198	-0.002	762	0.0 %	762				
2.5	2.500	2.441	-0.059	9562	0.0 %	9562				
5.0	5.000	4.997	-0.003	19591	0.0 %	19591				
7.5	7.500	7.653	0.153	30012	0.0 %	30012				
10.0	10.000	9.897	-0.103	38816	0.0 %	38816				

LB112474 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Dec 2020 14:27:28

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	40	-	0.00007470A	PPB	21 Dec 2020 14:46:41	S	Std
0.2 - 1	50.2	762	-	0.20007470A	PPB	21 Dec 2020 14:48:58	S	Std
2.5 - 1	52.5	9562	-	2.50007470A	PPB	21 Dec 2020 14:51:16	S	Std
5.0 - 1	35	19591	-	5.00007470A	PPB	21 Dec 2020 14:53:33	S	Std
7.5 - 1	57.5	30012	-	7.50007470A	PPB	21 Dec 2020 14:55:49	S	Std
10.0 - 1	50	38816	-	10.00007470A	PPB	21 Dec 2020 15:00:45	S	Std
ICV49 - 1	ICV49	14466	3.6908	-7470A	PPB	21 Dec 2020 15:04:36	U	SMPL
ICB49 - 1	ICB49	-24	-0.0027	-7470A	PPB	21 Dec 2020 15:06:51	U	SMPL
CCV15 - 1	CCV15	19850	5.0631	-7470A	PPB	21 Dec 2020 15:09:06	U	SMPL
CCB15 - 1	CCB15	-42	-0.0073	-7470A	PPB	21 Dec 2020 15:11:21	U	SMPL
CRA - 1	CRA	743	0.1928	-7470A	PPB	21 Dec 2020 15:13:36	U	SMPL
HighStd - 1	HighStd	38620	9.8475	-7470A	PPB	21 Dec 2020 15:15:51	U	SMPL
ChkStd - 1	ChkStd	27014	6.8892	-7470A	PPB	21 Dec 2020 15:18:07	U	SMPL
PB133729BL - 1	PBW	-84	-0.0180	-7470A	PPB	21 Dec 2020 15:20:23	U	SMPL
PB133729BS - 1	LCSW	15206	3.8794	-7470A	PPB	21 Dec 2020 15:22:39	U	SMPL
L5126-11 - 1	CONTAINMENT-DEBRIS	28	0.0106	-7470A	PPB	21 Dec 2020 15:24:56	U	SMPL
L5158-03 - 1	32841	34	0.0121	-7470A	PPB	21 Dec 2020 15:27:12	U	SMPL
L5159-02 - 1	CGP-1	23	0.0093	-7470A	PPB	21 Dec 2020 15:29:29	U	SMPL
L5159-04 - 1	CGP-2	36	0.0126	-7470A	PPB	21 Dec 2020 15:31:47	U	SMPL
L5159-06 - 1	CGP-3	70	0.0213	-7470A	PPB	21 Dec 2020 15:34:02	U	SMPL
CCV16 - 1	CCV16	19568	4.9912	-7470A	PPB	21 Dec 2020 15:36:18	U	SMPL
CCB16 - 1	CCB16	-34	-0.0052	-7470A	PPB	21 Dec 2020 15:38:33	U	SMPL
L5159-08 - 1	CGP-4	13	0.0067	-7470A	PPB	21 Dec 2020 15:40:49	U	SMPL
L5159-10 - 1	CGP-5	21	0.0088	-7470A	PPB	21 Dec 2020 15:43:04	U	SMPL
L5159-12 - 1	CGP-6	25	0.0098	-7470A	PPB	21 Dec 2020 15:45:20	U	SMPL
L5159-14 - 1	CGP-7	25	0.0098	-7470A	PPB	21 Dec 2020 15:47:35	U	SMPL
L5159-16 - 1	CGP-8	29	0.0108	-7470A	PPB	21 Dec 2020 15:49:51	U	SMPL
L5159-18 - 1	CGP-9	106	0.0304	-7470A	PPB	21 Dec 2020 15:52:07	U	SMPL
L5159-20 - 1	CGP-10	28	0.0106	-7470A	PPB	21 Dec 2020 15:54:24	U	SMPL
L5159-22 - 1	CGP-11	25	0.0098	-7470A	PPB	21 Dec 2020 15:56:40	U	SMPL
L5159-24 - 1	CGP-12	64	0.0197	-7470A	PPB	21 Dec 2020 15:58:58	U	SMPL
L5159-26 - 1	CGP-13	24	0.0095	-7470A	PPB	21 Dec 2020 16:01:13	U	SMPL
CCV17 - 1	CCV17	19450	4.9612	-7470A	PPB	21 Dec 2020 16:03:29	U	SMPL
CCB17 - 1	CCB17	-42	-0.0073	-7470A	PPB	21 Dec 2020 16:05:45	U	SMPL
L5159-28 - 1	CGP-14	18	0.0080	-7470A	PPB	21 Dec 2020 16:08:00	U	SMPL
L5159-28DUP - 1	CGP-14DUP	31	0.0113	-7470A	PPB	21 Dec 2020 16:10:16	U	SMPL
L5159-28MS - 1	CGP-14MS	14846	3.7876	-7470A	PPB	21 Dec 2020 16:12:31	U	SMPL
L5159-28MSD - 1	CGP-14MSD	14793	3.7741	-7470A	PPB	21 Dec 2020 16:14:47	U	SMPL
PB133729TB - 1	PB133702TB	-39	-0.0065	-7470A	PPB	21 Dec 2020 16:17:03	U	SMPL
L5159-28LX5 - 1		1	0.0037	-7470A	PPB	21 Dec 2020 16:19:19	U	SMPL
L5159-28A - 1		16199	4.1325	-7470A	PPB	21 Dec 2020 16:21:36	U	SMPL
CCV18 - 1	CCV18	19385	4.9446	-7470A	PPB	21 Dec 2020 16:23:53	U	SMPL
CCB18 - 1	CCB18	-48	-0.0088	-7470A	PPB	21 Dec 2020 16:26:10	U	SMPL