



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112493

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/22/2020	13:00
2	CAL2	1	Water	NA	NA	20.1	7.00	12/22/2020	13:01
3	CAL3	1	Water	NA	NA	20.1	10.03	12/22/2020	13:02
4	ICV	1	Water	NA	NA	20.2	7.01	12/22/2020	13:05
5	CCV1	1	Water	NA	NA	20.3	2.01	12/22/2020	13:08
6	L5194-01	1	Water	NA	NA	18.5	5.50	12/22/2020	13:10
7	L5194-01DUP	1	Water	NA	NA	18.6	5.51	12/22/2020	13:11
8	CCV2	1	Water	NA	NA	20.2	12.02	12/22/2020	13:15

JB 12493

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L5194 WorkList ID : 145174 Department : Wet-Chemistry Date : 12-22-2020 12:49:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/14/2020	Water	L5194-01	pH	Cool 4 deg C	STER17	M31	JS9600	12/14/2020	9040C

Date/Time 12-22-2020 12:50
Received by: JP [Signature]
Relinquished by: JP [Signature]

Date/Time 12-22-2020 14:00
Received by: [Signature]
Relinquished by: JP [Signature]