


LB112496

7471B

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	240	0.000	240				
0.2	0.200	0.213	0.013	1583	0.0 %	1583				
2.5	2.500	2.588	0.088	15310	0.0 %	15310				
5.0	5.000	5.043	0.043	29499	0.0 %	29499				
7.5	7.500	7.223	-0.277	42102	0.0 %	42102				
10.0	10.000	10.153	0.153	59039	0.0 %	59039				

LB112496 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Dec 2020 14:29:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	240	-	0.0000	7471B	PPB	22 Dec 2020 14:42:42	S	Std
0.2 - 1	50.2	1583	-	0.2000	7471B	PPB	22 Dec 2020 14:44:59	S	Std
2.5 - 1	52.5	15310	-	2.5000	7471B	PPB	22 Dec 2020 14:47:16	S	Std
5.0 - 1	55	29499	-	5.0000	7471B	PPB	22 Dec 2020 14:49:33	S	Std
7.5 - 1	57.5	42102	-	7.5000	7471B	PPB	22 Dec 2020 14:51:50	S	Std
10.0 - 1	510	59039	-	10.0000	7471B	PPB	22 Dec 2020 14:56:45	S	Std
ICV51 - 1	ICV51	22514	3.8341	-	7471B	PPB	22 Dec 2020 14:59:26	U	SMPL
ICB51 - 1	ICB51	211	-0.0245	-	7471B	PPB	22 Dec 2020 15:04:19	U	SMPL
CCV21 - 1	CCV21	29661	5.0706	-	7471B	PPB	22 Dec 2020 15:06:34	U	SMPL
CCB21 - 1	CCB21	177	-0.0304	-	7471B	PPB	22 Dec 2020 15:08:49	U	SMPL
CRA - 1	CRA	1592	0.2144	-	7471B	PPB	22 Dec 2020 15:11:03	U	SMPL
HighStd - 1	HighStd	58777	10.1079	-	7471B	PPB	22 Dec 2020 15:13:18	U	SMPL
ChkStd - 1	ChkStd	40390	6.9268	-	7471B	PPB	22 Dec 2020 15:15:34	U	SMPL
PB133745BL - 1	PBS	305	-0.0082	-	7471B	PPB	22 Dec 2020 15:17:51	U	SMPL
PB133745BS - 1	LCSS	21182	3.6037	-	7471B	PPB	22 Dec 2020 15:20:08	U	SMPL
L5155-01 - 1	TR-04-12-18-20	5041	0.8111	-	7471B	PPB	22 Dec 2020 15:22:25	U	SMPL
L5176-01 - 1	CHRT-26878	1794	0.2494	-	7471B	PPB	22 Dec 2020 15:24:41	U	SMPL
L5176-01DUP - 1	CHRT-26878DUP	1658	0.2259	-	7471B	PPB	22 Dec 2020 15:26:58	U	SMPL
L5176-01MS - 1	CHRT-26878MS	21240	3.6137	-	7471B	PPB	22 Dec 2020 15:29:14	U	SMPL
L5176-01MSD - 1	CHRT-26878MSD	21423	3.6454	-	7471B	PPB	22 Dec 2020 15:31:30	U	SMPL
CCV22 - 1	CCV22	29409	5.0270	-	7471B	PPB	22 Dec 2020 15:33:45	U	SMPL
CCB22 - 1	CCB22	237	-0.0200	-	7471B	PPB	22 Dec 2020 15:38:35	U	SMPL
L5176-03 - 1	CHRT-27339	2872	0.4359	-	7471B	PPB	22 Dec 2020 15:40:50	U	SMPL
L5176-05 - 1	CHRT-26920	1900	0.2677	-	7471B	PPB	22 Dec 2020 15:43:05	U	SMPL
L5176-07 - 1	CHRT-25299	3895	0.6129	-	7471B	PPB	22 Dec 2020 15:45:21	U	SMPL
L5179-01 - 1	CHRT-24743	663	0.0537	-	7471B	PPB	22 Dec 2020 15:47:37	U	SMPL
L5176-01LX5 - 1		848	0.0857	-	7471B	PPB	22 Dec 2020 15:52:47	U	SMPL
L5176-01A - 1		23604	4.0227	-	7471B	PPB	22 Dec 2020 15:55:03	U	SMPL
CCV23 - 1	CCV23	28964	4.9500	-	7471B	PPB	22 Dec 2020 15:57:20	U	SMPL
CCB23 - 1	CCB23	148	-0.0354	-	7471B	PPB	22 Dec 2020 15:59:37	U	SMPL