

LB112511


ISM02.4

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

43850



A= 0.0000e+000

B= 2.2836e-004

C= -9.7729e-003

Rho= 0.9999859

Accept=Accepted

slope  
4.1m/2m

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.00   | 0.000  | 0.003  | 0.003  | 54    | 0.000      | 54    |       |       |       |       | -4 |
| 0.05   | 0.050  | 0.048  | -0.002 | 253   | 0.0 %      | 253   |       |       |       |       | -3 |
| 0.20   | 0.200  | 0.195  | -0.005 | 897   | 0.0 %      | 897   |       |       |       |       | 0  |
| 2.50   | 2.500  | 2.488  | -0.012 | 10938 | 0.0 %      | 10938 |       |       |       |       | 1  |
| 5.00   | 5.000  | 5.041  | 0.041  | 22120 | 0.0 %      | 22120 |       |       |       |       | 0  |
| 7.50   | 7.500  | 7.471  | -0.029 | 32761 | 0.0 %      | 32761 |       |       |       |       | 0  |
| 10.0   | 10.000 | 10.004 | 0.004  | 43850 | 0.0 %      | 43850 |       |       |       |       | 0  |

LB112511  
INSTRUMENT ID :CV1

| Sample ID  | Extended ID  | $\mu$ Abs. | Conc.     | Std Conc | Method    | Units | Date             | Type |
|------------|--------------|------------|-----------|----------|-----------|-------|------------------|------|
|            | 0 S0.0       | 54         | -         |          | 0 ISM02.4 | PPB   | 12/23/2020 11:17 | Std  |
|            | 0.05 S0.05   | 253        | -         | 0.05     | ISM02.4   | PPB   | 12/23/2020 11:19 | Std  |
|            | 0.2 S0.2     | 897        | -         | 0.2      | ISM02.4   | PPB   | 12/23/2020 11:21 | Std  |
|            | 2.5 S2.5     | 10938      | -         | 2.5      | ISM02.4   | PPB   | 12/23/2020 11:24 | Std  |
|            | 5 S5.0       | 22120      | -         | 5        | ISM02.4   | PPB   | 12/23/2020 11:28 | Std  |
|            | 7.5 S7.5     | 32761      | -         | 7.5      | ISM02.4   | PPB   | 12/23/2020 11:31 | Std  |
|            | 10 S10.0     | 43850      | -         | 10       | ISM02.4   | PPB   | 12/23/2020 11:33 | Std  |
| ICV        | ICV          | 16160      | 3.6804 -  |          | ISM02.4   | PPB   | 12/23/2020 11:36 | SMPL |
| ICB        | ICB          | -15        | -0.0132 - |          | ISM02.4   | PPB   | 12/23/2020 11:38 | SMPL |
| CCV024     | CCV024       | 22086      | 5.0337 -  |          | ISM02.4   | PPB   | 12/23/2020 11:40 | SMPL |
| CCB024     | CCB024       | -20        | -0.0143 - |          | ISM02.4   | PPB   | 12/23/2020 11:42 | SMPL |
| PB133778BL | PBW778       | 42         | -0.0002 - |          | ISM02.4   | PPB   | 12/23/2020 11:45 | SMPL |
| L4485-08   | MDL-WATER-01 | 255        | 0.0485 -  |          | ISM02.4   | PPB   | 12/23/2020 11:47 | SMPL |
| L4485-09   | MDL-WATER-02 | 259        | 0.0494 -  |          | ISM02.4   | PPB   | 12/23/2020 11:49 | SMPL |
| L4485-10   | MDL-WATER-03 | 249        | 0.0471 -  |          | ISM02.4   | PPB   | 12/23/2020 11:51 | SMPL |
| L4485-11   | MDL-WATER-04 | 272        | 0.0523 -  |          | ISM02.4   | PPB   | 12/23/2020 11:54 | SMPL |
| L4485-12   | MDL-WATER-05 | 297        | 0.058 -   |          | ISM02.4   | PPB   | 12/23/2020 12:03 | SMPL |
| L4485-13   | MDL-WATER-06 | 307        | 0.0603 -  |          | ISM02.4   | PPB   | 12/23/2020 12:05 | SMPL |
| L4485-14   | MDL-WATER-07 | 306        | 0.0601 -  |          | ISM02.4   | PPB   | 12/23/2020 12:10 | SMPL |
| CCV025     | CCV025       | 21730      | 4.9524 -  |          | ISM02.4   | PPB   | 12/23/2020 12:12 | SMPL |
| CCB025     | CCB025       | -12        | -0.0125 - |          | ISM02.4   | PPB   | 12/23/2020 12:14 | SMPL |
| PB133781BL | PBW781       | 31         | -0.0027 - |          | ISM02.4   | PPB   | 12/23/2020 12:16 | SMPL |
| L5030-01   | MEW0T9       | 73         | 0.0069 -  |          | ISM02.4   | PPB   | 12/23/2020 12:19 | SMPL |
| L5030-02   | MEW0W1       | 43         | 0 -       |          | ISM02.4   | PPB   | 12/23/2020 12:21 | SMPL |
| L5030-03   | MEW0W5       | 41         | -0.0004 - |          | ISM02.4   | PPB   | 12/23/2020 12:23 | SMPL |
| L5030-04   | MEW0W5D      | 45         | 0.0005 -  |          | ISM02.4   | PPB   | 12/23/2020 12:25 | SMPL |
| L5030-05   | MEW0W5S      | 4416       | 0.9986 -  |          | ISM02.4   | PPB   | 12/23/2020 12:28 | SMPL |
| CCV026     | CCV026       | 21688      | 4.9428 -  |          | ISM02.4   | PPB   | 12/23/2020 12:30 | SMPL |
| CCB026     | CCB026       | -15        | -0.0132 - |          | ISM02.4   | PPB   | 12/23/2020 12:32 | SMPL |