



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/24/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 12/23/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 12/24/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB112515

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L5173-06	BG4R2	1	1.14	9.92	9.42	94.3	
L5173-07	BG4R3	2	1.19	9.7	8.47	85.5	
L5173-08	BG4R4	3	1.14	9.98	8.46	82.8	
L5173-09	BG4R5	4	1.15	9.8	8.17	81.2	
L5173-10	BG4R6	5	1.13	9.97	7.85	76.0	
L5173-11	BG4R7	6	1.18	9.62	7.41	73.8	
L5173-12	BG4R8	7	1.15	9.96	7.04	66.9	
L5173-13	VHBLK02	8	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

VB 112515

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-I5173 WorkList ID : 145202 Department : Wet-Chemistry Date : 12-23-2020 13:42:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/31/2020	Solid	L5173-06	Percent Solids	1:1 HCl to pH <	USEP04	A23	BG4R2	12/21/2020	Chemtech -SO
12/31/2020	Solid	L5173-07	Percent Solids	1:1 HCl to pH <	USEP04	A23	BG4R3	12/21/2020	Chemtech -SO
12/31/2020	Solid	L5173-08	Percent Solids	1:1 HCl to pH <	USEP04	A23	BG4R4	12/21/2020	Chemtech -SO
12/31/2020	Solid	L5173-09	Percent Solids	1:1 HCl to pH <	USEP04	A23	BG4R5	12/21/2020	Chemtech -SO
12/31/2020	Solid	L5173-10	Percent Solids	1:1 HCl to pH <	USEP04	A23	BG4R6	12/21/2020	Chemtech -SO
12/31/2020	Solid	L5173-11	Percent Solids	1:1 HCl to pH <	USEP04	A23	BG4R7	12/21/2020	Chemtech -SO
12/31/2020	Solid	L5173-12	Percent Solids	1:1 HCl to pH <	USEP04	A23	BG4R8	12/21/2020	Chemtech -SO
01/01/2021	Solid	L5173-13	Percent Solids	1:1 HCl to pH <	USEP04	A23	VHBLK02	12/22/2020	Chemtech -SO

Date/Time 12.23.2020 13:50
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 12.23.2020 15:00
 Received by: [Signature]
 Relinquished by: [Signature]