



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112528

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/24/2020	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	12/24/2020	11:11
3	CAL3	1	Water	NA	NA	20.2	10.03	12/24/2020	11:12
4	ICV	1	Water	NA	NA	20.2	7.01	12/24/2020	11:15
5	CCV1	1	Water	NA	NA	20.3	2.02	12/24/2020	11:16
6	L5215-03	1	Solid	20.02	20	19.1	5.80	12/24/2020	11:19
7	L5215-03DUP	1	Solid	20.03	20	19.2	5.81	12/24/2020	11:20
8	CCV2	1	Water	NA	NA	20.2	12.02	12/24/2020	11:22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion 15215 WorkList ID : 145236 Department : Wet-Chemistry Date : 12-24-2020 10:50:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/24/2020	Solid	L5215-03	Corrosivity	Cool 4 deg C	PSEG03	M51	DRUMS-1-4	12/24/2020	9045D

Date/Time 12-24-2020 11:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 12-24-2020 12:10:00
 Received by: [Signature]
 Relinquished by: [Signature]