

LB112529


ISM02.4

INSTRUMENT ID: CVA

Linear

 μ Abs.:

45447



A= 0.0000e+000

B= 2.2254e-004

C= 1.7899e-002

Rho= 0.9997018

Accept=Accepted

slope
y=1.7899e-002

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	0.022	0.022	20	0.000	20					-
0.05	0.050					253					10
0.20	0.200	0.219	0.019	903	0.0 %	903					2
2.50	2.500	2.544	0.044	11351	0.0 %	11351					-1
5.00	5.000	4.941	-0.059	22123	0.0 %	22123					-2
7.50	7.500	7.342	-0.158	32909	0.0 %	32909					1
10.0	10.000	10.132	0.132	45447	0.0 %	45447					

LB112529
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	20	-		0 ISM02.4	PPB	12/24/2020 11:08	Std
	0.2 S0.2	903	-		0.2 ISM02.4	PPB	12/24/2020 11:10	Std
	2.5 S2.5	11351	-		2.5 ISM02.4	PPB	12/24/2020 11:12	Std
	5 S5	22123	-		5 ISM02.4	PPB	12/24/2020 11:14	Std
	7.5 S7.5	32909	-		7.5 ISM02.4	PPB	12/24/2020 11:17	Std
	10 S10	45447	-		10 ISM02.4	PPB	12/24/2020 11:22	Std
ICV	ICV	16824	3.762 -		ISM02.4	PPB	12/24/2020 11:25	SMPL
ICB	ICB	-4	0.017 -		ISM02.4	PPB	12/24/2020 11:27	SMPL
CCV027	CCV027	22153	4.9479 -		ISM02.4	PPB	12/24/2020 11:29	SMPL
CCB027	CCB027	-17	0.0141 -		ISM02.4	PPB	12/24/2020 11:31	SMPL
PB133790BL	PBW790	90	0.0379 -		ISM02.4	PPB	12/24/2020 11:34	SMPL
L5187-01	MBG4R8	70	0.0335 -		ISM02.4	PPB	12/24/2020 11:36	SMPL
L5187-02	MBG4R9	78	0.0353 -		ISM02.4	PPB	12/24/2020 11:38	SMPL
L5187-03	MBG4S0	157	0.0528 -		ISM02.4	PPB	12/24/2020 11:40	SMPL
L5187-04	MBG4S2	102	0.0406 -		ISM02.4	PPB	12/24/2020 11:43	SMPL
L5187-05	MBG4S4	115	0.0435 -		ISM02.4	PPB	12/24/2020 11:45	SMPL
L5187-06	MBG4S5	124	0.0455 -		ISM02.4	PPB	12/24/2020 11:47	SMPL
L5187-07	MBG4S1	126	0.0459 -		ISM02.4	PPB	12/24/2020 11:50	SMPL
L5187-08	MBG4S6	85	0.0368 -		ISM02.4	PPB	12/24/2020 11:52	SMPL
L5187-09	MBG4S6D	99	0.0399 -		ISM02.4	PPB	12/24/2020 11:54	SMPL
L5187-10	MBG4S6S	4154	0.9424 -		ISM02.4	PPB	12/24/2020 11:56	SMPL
L5187-11	MBG4S7	70	0.0335 -		ISM02.4	PPB	12/24/2020 11:59	SMPL
L5187-12	MBG4S9	73	0.0341 -		ISM02.4	PPB	12/24/2020 12:01	SMPL
L5187-13	MBG4T0	71	0.0337 -		ISM02.4	PPB	12/24/2020 12:03	SMPL
L5187-14	MBG4T1	90	0.0379 -		ISM02.4	PPB	12/24/2020 12:05	SMPL
L5187-15	MBG4T2	87	0.0373 -		ISM02.4	PPB	12/24/2020 12:08	SMPL
PB133791BL	PBW791	75	0.0346 -		ISM02.4	PPB	12/24/2020 12:10	SMPL
L5191-01	MBG4Z2	80	0.0357 -		ISM02.4	PPB	12/24/2020 12:12	SMPL
L5191-02	MBG4Z3	79	0.0355 -		ISM02.4	PPB	12/24/2020 12:14	SMPL
L5191-03	MBG4Z4	93	0.0386 -		ISM02.4	PPB	12/24/2020 12:17	SMPL
CCV028	CCV028	22072	4.9299 -		ISM02.4	PPB	12/24/2020 12:19	SMPL
CCB028	CCB028	-2	0.0175 -		ISM02.4	PPB	12/24/2020 12:21	SMPL
L5191-04	MBG4Z5	42	0.0272 -		ISM02.4	PPB	12/24/2020 12:24	SMPL
L5191-05	MBG4Z6	42	0.0272 -		ISM02.4	PPB	12/24/2020 12:26	SMPL
L5191-06	MBG4Z7	59	0.031 -		ISM02.4	PPB	12/24/2020 12:28	SMPL
L5191-07	MBG4T3	55	0.0301 -		ISM02.4	PPB	12/24/2020 12:30	SMPL
L5191-08	MBG4T4	61	0.0315 -		ISM02.4	PPB	12/24/2020 12:33	SMPL
L5191-09	MBG4T5	64	0.0321 -		ISM02.4	PPB	12/24/2020 12:35	SMPL
L5191-10	MBG4T5D	103	0.0408 -		ISM02.4	PPB	12/24/2020 12:37	SMPL
L5191-11	MBG4T5S	4315	0.9782 -		ISM02.4	PPB	12/24/2020 12:39	SMPL
L5191-12	MBG4T6	39	0.0266 -		ISM02.4	PPB	12/24/2020 12:42	SMPL
L5191-13	MBG4T7	57	0.0306 -		ISM02.4	PPB	12/24/2020 12:44	SMPL
L5191-14	MBG4T8	57	0.0306 -		ISM02.4	PPB	12/24/2020 12:46	SMPL
L5191-15	MBG4T9	55	0.0301 -		ISM02.4	PPB	12/24/2020 12:48	SMPL
CCV029	CCV029	22107	4.9377 -		ISM02.4	PPB	12/24/2020 12:51	SMPL
CCB029	CCB029	2	0.0183 -		ISM02.4	PPB	12/24/2020 12:53	SMPL