


LB112545

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

37636



Linear

A= 0.0000e+000

B= 2.6554e-004

C= -1.8744e-002

Rho= 0.9999728

Accept=Accepted

slope
 $y = mx + \text{intercept}$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.012	-0.012	26	0.000	26					2
0.05	0.050	0.051	0.001	262	0.0 %	262					12
0.20	0.200	0.223	0.023	912	0.0 %	912					-2
2.50	2.500	2.454	-0.046	9311	0.0 %	9311					1
5.00	5.000	5.034	0.034	19030	0.0 %	19030					0
7.50	7.500	7.525	0.025	28408	0.0 %	28408					0
10.0	10.000	9.975	-0.025	37636	0.0 %	37636					0

LB112545

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0	S0.05	26	-		0 SFAM01.1	PPB	12/28/2020 13:49	Std
0.05	S0.05	262	-		0.05 SFAM01.1	PPB	12/28/2020 13:52	Std
0.2	S0.2	912	-		0.2 SFAM01.1	PPB	12/28/2020 13:54	Std
2.5	S2.5	9311	-		2.5 SFAM01.1	PPB	12/28/2020 13:56	Std
5	S5.0	19030	-		5 SFAM01.1	PPB	12/28/2020 13:59	Std
7.5	S7.5	28408	-		7.5 SFAM01.1	PPB	12/28/2020 14:01	Std
10	S10.0	37636	-		10 SFAM01.1	PPB	12/28/2020 14:06	Std
ICV	ICV	14350	3.7917	-	SFAM01.1	PPB	12/28/2020 14:11	SMPL
ICB	ICB	51	-0.0052	-	SFAM01.1	PPB	12/28/2020 14:16	SMPL
CCV037	CCV037	18908	5.002	-	SFAM01.1	PPB	12/28/2020 14:18	SMPL
CCB037	CCB037	51	-0.0052	-	SFAM01.1	PPB	12/28/2020 14:23	SMPL
PBW008	PBW008	37	-0.0089	-	SFAM01.1	PPB	12/28/2020 14:25	SMPL
PBW009	PBW009	49	-0.0057	-	SFAM01.1	PPB	12/28/2020 14:28	SMPL
PBW010	PBW010	61	-0.0025	-	SFAM01.1	PPB	12/28/2020 14:30	SMPL
L4485-08	MDL-WATER-01	264	0.0514	-	SFAM01.1	PPB	12/28/2020 14:32	SMPL
L4485-09	MDL-WATER-02	229	0.0421	-	SFAM01.1	PPB	12/28/2020 14:34	SMPL
L4485-10	MDL-WATER-03	257	0.0495	-	SFAM01.1	PPB	12/28/2020 14:37	SMPL
CCV038	CCV038	18768	4.9648	-	SFAM01.1	PPB	12/28/2020 14:39	SMPL
CCB038	CCB038	41	-0.0079	-	SFAM01.1	PPB	12/28/2020 14:44	SMPL