



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112554

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	12/29/2020	12:50
2	L5223-03		1	9	100.00	>212.0	>212.0	12/29/2020	13:20
3	L5223-03DUP		1	9	100.00	>212.0	>212.0	12/29/2020	13:45
4	L5223-07		1	10	100.00	>212.0	>212.0	12/29/2020	14:10
5	L5224-05		1	11	100.00	>212.0	>212.0	12/29/2020	14:35
6	L5226-02		1	9	100.00	>212.0	>212.0	12/29/2020	15:05
7	L5226-03		1	10	100.00	>212.0	>212.0	12/29/2020	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 112554

WorkList Name : fp122920 WorkList ID : 145289 Department : Wet-Chemistry Date : 12-29-2020 09:49:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2021	Water	L5223-03	Flash Point	Cool 4 deg C	PSEG03	L51	TOTE-COMP	12/28/2020	1010A
01/11/2021	Water	L5223-07	Flash Point	Cool 4 deg C	PSEG03	L51	DRUM-COMP	12/28/2020	1010A
01/11/2021	Water	L5224-05	Flash Point	Cool 4 deg C	PSEG03	L51	DRUM-COMP-2	12/28/2020	1010A
01/04/2021	Water	L5226-02	Flash Point	Cool 4 deg C	PSEG03	L51	16882	12/28/2020	1010A
01/04/2021	Water	L5226-03	Flash Point	Cool 4 deg C	PSEG03	L51	16883	12/28/2020	1010A

Date/Time 12/29/20 12:30
 Received by: KS (Weber)
 Relinquished by: AP CEXT. Lab?

Date/Time 12/29/20 15:40
 Received by: AP CEXT. Lab?
 Relinquished by: KS (Weber)