



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112555

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/29/2020	10:45
2	CAL2	1	Water	NA	NA	20.2	7.00	12/29/2020	10:46
3	CAL3	1	Water	NA	NA	20.2	10.03	12/29/2020	10:47
4	ICV	1	Water	NA	NA	20.2	7.00	12/29/2020	10:48
5	CCV1	1	Water	NA	NA	20.2	2.01	12/29/2020	10:50
6	L5223-05	1	Water	NA	NA	15.7	5.43	12/29/2020	11:00
7	L5223-05DUP	1	Water	NA	NA	15.8	5.44	12/29/2020	11:01
8	L5223-06	1	Water	NA	NA	15.3	7.54	12/29/2020	11:03
9	L5223-09	1	Water	NA	NA	16.0	5.63	12/29/2020	11:05
10	L5223-10	1	Water	NA	NA	16.6	5.48	12/29/2020	11:06
11	L5224-07	1	Water	NA	NA	16.5	6.66	12/29/2020	11:10
12	L5224-08	1	Water	NA	NA	17.0	7.55	12/29/2020	11:13
13	L5226-02	1	Water	NA	NA	19.9	5.48	12/29/2020	11:15
14	L5226-03	1	Water	NA	NA	19.4	7.33	12/29/2020	11:18
15	CCV2	1	Water	NA	NA	20.3	12.02	12/29/2020	11:20

WORKLIST(Hardcopy Internal Chain)

1212555

WorkList Name : ph 15200 WorkList ID : 145290 Department : Wet-Chemistry Date : 12-29-2020 10:37:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/28/2020	Water	L5223-05	pH	Cool 4 deg C	PSEG03	L51	31311	12/28/2020	9040C
12/28/2020	Water	L5223-06	pH	Cool 4 deg C	PSEG03	L51	31356	12/28/2020	9040C
12/28/2020	Water	L5223-09	pH	Cool 4 deg C	PSEG03	L51	31352	12/28/2020	9040C
12/28/2020	Water	L5223-10	pH	Cool 4 deg C	PSEG03	L51	31361	12/28/2020	9040C
12/28/2020	Water	L5224-07	pH	Cool 4 deg C	PSEG03	L51	31354	12/28/2020	9040C
12/28/2020	Water	L5224-08	pH	Cool 4 deg C	PSEG03	L51	31355	12/28/2020	9040C
12/28/2020	Water	L5226-02	pH	Cool 4 deg C	PSEG03	L51	16882	12/28/2020	9040C
12/28/2020	Water	L5226-03	pH	Cool 4 deg C	PSEG03	L51	16883	12/28/2020	9040C

Date/Time 12-29-2020 10:40
 Received by: JP (SM)
 Relinquished by: JP (SM)

Date/Time 12-29-2020 12:35
 Received by: JP (SM)
 Relinquished by: JP