



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112556

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/29/2020	10:45
2	CAL2	1	Water	NA	NA	20.2	7.00	12/29/2020	10:46
3	CAL3	1	Water	NA	NA	20.2	10.03	12/29/2020	10:47
4	ICV	1	Water	NA	NA	20.2	7.00	12/29/2020	10:48
5	CCV1	1	Water	NA	NA	20.2	2.01	12/29/2020	10:50
6	L5222-01	1	Solid	20.03	20	23.0	6.71	12/29/2020	11:25
7	L5222-01DUP	1	Solid	20.04	20	23.1	6.72	12/29/2020	11:26
8	L5222-03	1	Solid	20.01	20	20.6	10.15	12/29/2020	11:27
9	L5222-05	1	Solid	20.03	20	23.4	6.87	12/29/2020	11:29
10	L5222-07	1	Solid	20.03	20	20.5	9.74	12/29/2020	11:30
11	L5222-09	1	Solid	20.02	20	20.9	9.64	12/29/2020	11:32
12	L5222-11	1	Solid	20.03	20	19.9	10.39	12/29/2020	11:35
13	L5222-13	1	Solid	20.01	20	20.3	7.72	12/29/2020	11:36
14	L5222-15	1	Solid	20.03	20	20.2	10.11	12/29/2020	11:39
15	L5222-17	1	Solid	20.03	20	19.5	9.81	12/29/2020	11:40
16	CCV2	1	Water	NA	NA	20.2	12.02	12/29/2020	11:42
17	L5222-19	1	Solid	20.04	20	19.8	10.12	12/29/2020	11:45
18	L5223-01	1	Solid	20.03	20	19.4	7.26	12/29/2020	11:46
19	L5224-03	1	Solid	20.04	20	19.6	6.23	12/29/2020	11:48
20	L5224-04	1	Solid	20.02	20	19.4	6.13	12/29/2020	11:49

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.2	2.02	12/29/2020	11:50

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s I5200 WorkList ID : 145291 Department : Wet-Chemistry Date : 12-29-2020 10:38:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/28/2020	Solid	L5222-01	pH	Cool 4 deg C	PSEG03	L51	31371	12/28/2020	9045D
12/28/2020	Solid	L5222-03	pH	Cool 4 deg C	PSEG03	L51	31370	12/28/2020	9045D
12/28/2020	Solid	L5222-05	pH	Cool 4 deg C	PSEG03	L51	31369	12/28/2020	9045D
12/28/2020	Solid	L5222-07	pH	Cool 4 deg C	PSEG03	L51	31368	12/28/2020	9045D
12/28/2020	Solid	L5222-09	pH	Cool 4 deg C	PSEG03	L51	31367	12/28/2020	9045D
12/28/2020	Solid	L5222-11	pH	Cool 4 deg C	PSEG03	L51	31366	12/28/2020	9045D
12/28/2020	Solid	L5222-13	pH	Cool 4 deg C	PSEG03	L51	31365	12/28/2020	9045D
12/28/2020	Solid	L5222-15	pH	Cool 4 deg C	PSEG03	L51	31364	12/28/2020	9045D
12/28/2020	Solid	L5222-17	pH	Cool 4 deg C	PSEG03	L51	31363	12/28/2020	9045D
12/28/2020	Solid	L5222-19	pH	Cool 4 deg C	PSEG03	L51	31362	12/28/2020	9045D
12/28/2020	Solid	L5223-01	pH	Cool 4 deg C	PSEG03	L51	31303	12/28/2020	9045D
12/28/2020	Solid	L5224-03	pH	Cool 4 deg C	PSEG03	L51	TRE-1132	12/28/2020	9045D
12/28/2020	Solid	L5224-04	pH	Cool 4 deg C	PSEG03	L51	TRE-1133	12/28/2020	9045D

Date/Time 12-29-2020 10:40
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 12-29-2020 12:40
 Received by: [Signature]
 Relinquished by: [Signature]