


LB112560

SFAM01.1

INSTRUMENT ID: CV2

 μ Abs.:

44833



Linear

A= 0.0000e+000

B= 2.2532e-004

C= -4.1097e-003

Rho= 0.9997838

Accept=Accepted

Handwritten notes:
 slope
 y = Bx + intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.001	-0.001	16	0.000	16					
0.05	0.050					261					
0.20	0.200	0.235	0.035	1063	0.0 %	1063					
2.50	2.500	2.490	-0.010	11068	0.0 %	11068					
5.00	5.000	5.031	0.031	22344	0.0 %	22344					
7.50	7.500	7.347	-0.153	32624	0.0 %	32624					
10.0	10.000	10.098	0.098	44833	0.0 %	44833					

Handwritten notes:
 18
 0
 1
 -2
 1

LB112560

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	16	-		0 SFAM01.1	PPB	12/29/2020 14:49	Std
	0.2 S01	1063	-		0.2 SFAM01.1	PPB	12/29/2020 14:54	Std
	2.5 S02	11068	-		2.5 SFAM01.1	PPB	12/29/2020 14:56	Std
	5 S03	22344	-		5 SFAM01.1	PPB	12/29/2020 14:58	Std
	7.5 S04	32624	-		7.5 SFAM01.1	PPB	12/29/2020 15:00	Std
	10 S05	44833	-		10 SFAM01.1	PPB	12/29/2020 15:03	Std
ICV001	ICV001	17096	3.848 -		SFAM01.1	PPB	12/29/2020 15:08	SMPL
ICB001	ICB001	-32	-0.0113 -		SFAM01.1	PPB	12/29/2020 15:10	SMPL
CCV044	CCV044	21964	4.9449 -		SFAM01.1	PPB	12/29/2020 15:12	SMPL
CCB044	CCB044	-37	-0.0124 -		SFAM01.1	PPB	12/29/2020 15:15	SMPL
PB133845BL	PBW845	23	0.0011 -		SFAM01.1	PPB	12/29/2020 15:17	SMPL
L5181-07	X2T32	18086	4.0711 -		SFAM01.1	PPB	12/29/2020 15:21	SMPL
L5181-08	X2T32D	18070	4.0675 -		SFAM01.1	PPB	12/29/2020 15:26	SMPL
L5181-09	X2T32S	22449	5.0542 -		SFAM01.1	PPB	12/29/2020 15:33	SMPL
CCV045	CCV045	21967	4.9456 -		SFAM01.1	PPB	12/29/2020 15:35	SMPL
CCB045	CCB045	-45	-0.0142 -		SFAM01.1	PPB	12/29/2020 15:38	SMPL