


LB 112563

7471B

INSTRUMENT ID: CVJ



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	32	0.000	32				
0.2	0.200	0.205	0.005	1388	0.0 %	1388				
2.5	2.500	2.458	-0.042	13699	0.0 %	13699				
5.0	5.000	5.163	0.163	28485	0.0 %	28485				
7.5	7.500	7.468	-0.032	41084	0.0 %	41084				
10.0	10.000	9.950	-0.050	54647	0.0 %	54647				

LB112563 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Dec 2020 16:14:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.00007471B	PPB		29 Dec 2020 16:15:47	S	Std
0.2 - 1	51.2	1388	-	0.20007471B	PPB		29 Dec 2020 16:18:03	S	Std
2.5 - 1	52.5	13699	-	2.50007471B	PPB		29 Dec 2020 16:20:20	S	Std
5.0 - 1	55	28485	-	5.00007471B	PPB		29 Dec 2020 16:22:37	S	Std
7.5 - 1	57.5	41084	-	7.50007471B	PPB		29 Dec 2020 16:24:54	S	Std
10.0 - 1	510	54647	-	10.00007471B	PPB		29 Dec 2020 16:27:10	S	Std
ICV55 - 1	ICV55	21790	3.9379	-7471B	PPB		29 Dec 2020 16:32:29	U	SMPL
ICB55 - 1	ICB55	-53	-0.0587	-7471B	PPB		29 Dec 2020 16:34:43	U	SMPL
CCV46 - 1	CCV46	28589	5.1819	-7471B	PPB		29 Dec 2020 16:36:59	U	SMPL
CCB46 - 1	CCB46	-57	-0.0594	-7471B	PPB		29 Dec 2020 16:39:14	U	SMPL
CRA - 1	CRA	1367	0.2012	-7471B	PPB		29 Dec 2020 16:41:29	U	SMPL
HighStd - 1	HighStd	55490	10.1039	-7471B	PPB		29 Dec 2020 16:43:44	U	SMPL
ChkStd - 1	ChkStd	39127	7.1100	-7471B	PPB		29 Dec 2020 16:46:00	U	SMPL
PB133846BL - 1	pbs	-97	-0.0667	-7471B	PPB		29 Dec 2020 16:48:17	U	SMPL
PB133846BS - 1	lcss	20343	3.6731	-7471B	PPB		29 Dec 2020 16:53:23	U	SMPL
L5221-02 - 1	124021	2071	0.3300	-7471B	PPB		29 Dec 2020 16:55:39	U	SMPL
L5221-02DUP - 1	124021DUP	2168	0.3477	-7471B	PPB		29 Dec 2020 17:00:38	U	SMPL
L5221-02MS - 1	124021MS	21008	3.7948	-7471B	PPB		29 Dec 2020 17:05:35	U	SMPL
L5221-02MSD - 1	124021MSD	21055	3.8034	-7471B	PPB		29 Dec 2020 17:07:52	U	SMPL
L5221-04 - 1	RBR-58	3219	0.5400	-7471B	PPB		29 Dec 2020 17:10:07	U	SMPL
CCV47 - 1	CCV47	27477	4.9784	-7471B	PPB		29 Dec 2020 17:12:24	U	SMPL
CCB47 - 1	CCB47	-51	-0.0583	-7471B	PPB		29 Dec 2020 17:14:40	U	SMPL
L5221-06 - 1	RBR-69	3407	0.5744	-7471B	PPB		29 Dec 2020 17:16:56	U	SMPL
L5223-01 - 1	31303	1837	0.2872	-7471B	PPB		29 Dec 2020 17:19:11	U	SMPL
L5224-01 - 1	DRUM-COMP-1	142	-0.0230	-7471B	PPB		29 Dec 2020 17:21:27	U	SMPL
L5221-021x5 - 1		425	0.0288	-7471B	PPB		29 Dec 2020 17:23:42	U	SMPL
L5221-02a - 1		26710	4.8381	-7471B	PPB		29 Dec 2020 17:25:58	U	SMPL
CCV48 - 1	CCV48	28906	5.2399	-7471B	PPB		29 Dec 2020 17:28:15	U	SMPL
CCB48 - 1	CCB48	-62	-0.0603	-7471B	PPB		29 Dec 2020 17:30:32	U	SMPL