


LB112572

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 3.0096e-004

C= -1.8392e-002

Rho= 0.9999718

Accept=Accepted

*slope
y = 1/n(x rep)*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.012	-0.012	22	0.000	22					14
0.05	0.050	0.057	0.007	250	0.0 %	250					15
0.20	0.200	0.230	0.030	826	0.0 %	826					0
2.50	2.500	2.508	0.008	8396	0.0 %	8396					0
5.00	5.000	4.943	-0.057	16486	0.0 %	16486					-1
7.50	7.500	7.492	-0.008	24955	0.0 %	24955					0
10.0	10.000	10.031	0.031	33392	0.0 %	33392					0

LB112572
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	22	-		0 SFAM01.1	PPB	12/30/2020 11:36	Std
	0.05 S0.05	250	-		0.05 SFAM01.1	PPB	12/30/2020 11:38	Std
	0.2 S0.2	826	-		0.2 SFAM01.1	PPB	12/30/2020 11:41	Std
	2.5 S2.5	8396	-		2.5 SFAM01.1	PPB	12/30/2020 11:43	Std
	5 S5.0	16486	-		5 SFAM01.1	PPB	12/30/2020 11:45	Std
	7.5 S7.5	24955	-		7.5 SFAM01.1	PPB	12/30/2020 11:47	Std
	10 S10.0	33392	-		10 SFAM01.1	PPB	12/30/2020 11:50	Std
ICV	ICV	13041	3.9064 -		SFAM01.1	PPB	12/30/2020 11:55	SMPL
ICB	ICB	42	-0.0058 -		SFAM01.1	PPB	12/30/2020 12:00	SMPL
CCV049	CCV049	16496	4.9462 -		SFAM01.1	PPB	12/30/2020 12:02	SMPL
CCB049	CCB049	-5	-0.0199 -		SFAM01.1	PPB	12/30/2020 12:04	SMPL
PBW013	PBW013	48	-0.0039 -		SFAM01.1	PPB	12/30/2020 12:07	SMPL
PBW014	PBW014	67	0.0018 -		SFAM01.1	PPB	12/30/2020 12:09	SMPL
L4485-13	MDL-WATER-06	265	0.0614 -		SFAM01.1	PPB	12/30/2020 12:14	SMPL
L4485-14	MDL-WATER-07	250	0.0568 -		SFAM01.1	PPB	12/30/2020 12:16	SMPL
CCV050	CCV050	16659	4.9952 -		SFAM01.1	PPB	12/30/2020 12:21	SMPL
CCB050	CCB050	64	0.0009 -		SFAM01.1	PPB	12/30/2020 12:26	SMPL