

LB112581


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44536



A= 0.0000e+000

B= 2.2436e-004

C= -2.9663e-002

Rho= 0.9999434

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.002	0.002	142	0.000	142					-
0.05	0.050					250					-11
0.20	0.200	0.179	-0.021	931	0.0 %	931					-1
2.50	2.500	2.470	-0.030	11141	0.0 %	11141					2
5.00	5.000	5.080	0.080	22776	0.0 %	22776					0
7.50	7.500	7.506	0.006	33586	0.0 %	33586					0
10.0	10.000	9.963	-0.037	44536	0.0 %	44536					0

LB112581
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	142	-		0 SFAM01.1	PPB	12/30/2020 16:12	Std
	0.2 S01	931	-		0.2 SFAM01.1	PPB	12/30/2020 16:14	Std
	2.5 S02	11141	-		2.5 SFAM01.1	PPB	12/30/2020 16:16	Std
	5 S03	22776	-		5 SFAM01.1	PPB	12/30/2020 16:19	Std
	7.5 S04	33586	-		7.5 SFAM01.1	PPB	12/30/2020 16:21	Std
	10 S05	44536	-		10 SFAM01.1	PPB	12/30/2020 16:23	Std
ICV002	ICV002	17693	3.94 -		SFAM01.1	PPB	12/30/2020 16:26	SMPL
ICB002	ICB002	33	-0.0223 -		SFAM01.1	PPB	12/30/2020 16:28	SMPL
CCV056	CCV056	22923	5.1134 -		SFAM01.1	PPB	12/30/2020 16:30	SMPL
CCB056	CCB056	22	-0.0247 -		SFAM01.1	PPB	12/30/2020 16:33	SMPL
PB133844BL	PBS844	59	-0.0164 -		SFAM01.1	PPB	12/30/2020 16:35	SMPL
L5181-19	X2T36	18079	4.0266 -		SFAM01.1	PPB	12/30/2020 16:37	SMPL
L5181-20	X2T36D	17993	4.0073 -		SFAM01.1	PPB	12/30/2020 16:42	SMPL
L5181-21	X2T36S	29197	6.521 -		SFAM01.1	PPB	12/30/2020 16:49	SMPL
CCV057	CCV057	22994	5.1293 -		SFAM01.1	PPB	12/30/2020 16:51	SMPL
CCB057	CCB057	9	-0.0276 -		SFAM01.1	PPB	12/30/2020 16:53	SMPL