


7470A

LB 112597
INSTRUMENT ID: CV4

Linear

 μ Abs.:

39941



A= 0.0000e+000

B= 2.5008e-004

C= -2.3737e-002

Rho= 0.9998673

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	153	0.000	153				
0.2	0.200	0.215	0.015	954	0.0 %	954				
2.5	2.500	2.489	-0.011	10046	0.0 %	10046				
5.0	5.000	4.909	-0.091	19726	0.0 %	19726				
7.5	7.500	7.608	0.108	30517	0.0 %	30517				
10.0	10.000	9.965	-0.035	39941	0.0 %	39941				

LB112597 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 31 Dec 2020 15:09:37				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	153	-	0.00007470A	PPB	31 Dec 2020 15:15:55	S	Std
0.2 - 1	50.2	954	-	0.20007470A	PPB	31 Dec 2020 15:18:12	S	Std
2.5 - 1	52.5	10046	-	2.50007470A	PPB	31 Dec 2020 15:20:29	S	Std
5.0 - 1	55	19726	-	5.00007470A	PPB	31 Dec 2020 15:22:46	S	Std
7.5 - 1	57.5	30517	-	7.50007470A	PPB	31 Dec 2020 15:25:03	S	Std
10.0 - 1	510	39941	-	10.00007470A	PPB	31 Dec 2020 15:27:19	S	Std
ICV58 - 1	ICV58	16598	4.1271	-7470A	PPB	31 Dec 2020 15:34:57	U	SMPL
ICB58 - 1	ICB58	38	-0.0142	-7470A	PPB	31 Dec 2020 15:37:12	U	SMPL
CCV61 - 1	CCV61	19613	4.8811	-7470A	PPB	31 Dec 2020 15:39:27	U	SMPL
CCB61 - 1	CCB61	97	0.0005	-7470A	PPB	31 Dec 2020 15:44:31	U	SMPL
CRA - 1	CRA	967	0.2181	-7470A	PPB	31 Dec 2020 15:46:46	U	SMPL
HighStd - 1	HighStd	40238	10.0390	-7470A	PPB	31 Dec 2020 15:49:00	U	SMPL
ChkStd - 1	ChkStd	27542	6.8640	-7470A	PPB	31 Dec 2020 15:51:16	U	SMPL
PB133884BL - 1	PBW	-20	-0.0287	-7470A	PPB	31 Dec 2020 15:53:32	U	SMPL
PB133884BS - 1	LCSW	13887	3.4491	-7470A	PPB	31 Dec 2020 15:55:48	U	SMPL
L5241-01 - 1	J6-OWS-5	25	-0.0175	-7470A	PPB	31 Dec 2020 15:58:04	U	SMPL
L5241-02 - 1	J6-OWS-4	277	0.0455	-7470A	PPB	31 Dec 2020 16:00:21	U	SMPL
L5241-03 - 1	J6-OWS-2	106	0.0028	-7470A	PPB	31 Dec 2020 16:02:38	U	SMPL
L5241-04 - 1	J6-OWS-1	173	0.0195	-7470A	PPB	31 Dec 2020 16:04:56	U	SMPL
L5241-05 - 1	J6-OWS-7	172	0.0193	-7470A	PPB	31 Dec 2020 16:07:11	U	SMPL
CCV62 - 1	CCV62	19727	4.9096	-7470A	PPB	31 Dec 2020 16:09:26	U	SMPL
CCB62 - 1	CCB62	53	-0.0105	-7470A	PPB	31 Dec 2020 16:11:42	U	SMPL
L5241-06 - 1	J6-OWS-ENGINE-WASH	93	-0.0005	-7470A	PPB	31 Dec 2020 16:13:58	U	SMPL
L5241-07 - 1	J6-OWS-BUS-WASH	95	0.0000	-7470A	PPB	31 Dec 2020 16:16:14	U	SMPL
L5242-15 - 1	SAMPLE-1(494+26)	97	0.0005	-7470A	PPB	31 Dec 2020 16:18:30	U	SMPL
L5242-16 - 1	SAMPLE-2(484+26)	109	0.0035	-7470A	PPB	31 Dec 2020 16:20:46	U	SMPL
L5242-17 - 1	SAMPLE-3(474+26)	98	0.0008	-7470A	PPB	31 Dec 2020 16:23:02	U	SMPL
L5242-18 - 1	SAMPLE-4(464+26)	92	-0.0007	-7470A	PPB	31 Dec 2020 16:25:19	U	SMPL
L5242-19 - 1	SAMPLE-5(454+26)	109	0.0035	-7470A	PPB	31 Dec 2020 16:27:36	U	SMPL
L5242-20 - 1	SAMPLE-6(444+26)	83	-0.0030	-7470A	PPB	31 Dec 2020 16:29:54	U	SMPL
L5242-21 - 1	SAMPLE-7(434+26)	92	-0.0007	-7470A	PPB	31 Dec 2020 16:32:11	U	SMPL
L5242-21DUP - 1	SAMPLE-7(434+26)DUP	92	-0.0007	-7470A	PPB	31 Dec 2020 16:34:28	U	SMPL
CCV63 - 1	CCV63	19752	4.9159	-7470A	PPB	31 Dec 2020 16:36:44	U	SMPL
CCB63 - 1	CCB63	41	-0.0135	-7470A	PPB	31 Dec 2020 16:39:00	U	SMPL
L5242-21MS - 1	SAMPLE-7(434+26)MS	14176	3.5214	-7470A	PPB	31 Dec 2020 16:41:17	U	SMPL
L5242-21MSD - 1	SAMPLE-7(434+26)MSD	13475	3.3461	-7470A	PPB	31 Dec 2020 16:43:33	U	SMPL
PB133884TB - 1	PB133874TB	49	-0.0115	-7470A	PPB	31 Dec 2020 16:45:50	U	SMPL
L5242-21LX5 - 1		100	0.0013	-7470A	PPB	31 Dec 2020 16:48:07	U	SMPL
L5242-21A - 1		16056	3.9916	-7470A	PPB	31 Dec 2020 16:50:24	U	SMPL
CCV64 - 1	CCV64	19574	4.8713	-7470A	PPB	31 Dec 2020 16:52:40	U	SMPL
CCB64 - 1	CCB64	36	-0.0147	-7470A	PPB	31 Dec 2020 16:54:58	U	SMPL