



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112610

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/04/2021	08:10
2	CAL2	1	Water	NA	NA	20.2	7.01	01/04/2021	08:11
3	CAL3	1	Water	NA	NA	20.3	10.04	01/04/2021	08:12
4	ICV	1	Water	NA	NA	20.3	7.00	01/04/2021	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	01/04/2021	08:16
6	L5248-04	1	Solid	20.03	20	20.1	7.54	01/04/2021	08:25
7	L5248-04DUP	1	Solid	20.02	20	20.2	7.55	01/04/2021	08:26
8	CCV2	1	Water	NA	NA	20.2	12.03	01/04/2021	08:30

WORKLIST(Hardcopy Internal Chain)

LB112610

WorkList Name : corrosivity I5148

WorkList ID : 145350

Department : Wet-Chemistry

Date : 01-04-2021 08:12:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/31/2020	Solid	L-5248-04	✓ Corrosivity	Cool 4 deg C	PSEG03	L61	VR16125	12/31/2020	9045D

Date/Time 01-04-21 8:05

Received by: JO Wet Chem

Relinquished by: SP (SM)

Date/Time 01-04-21 10:00

Received by: SP (SM)

Relinquished by: JO Wet Chem