

LB112613


7471B

INSTRUMENT ID: CVD

Linear

 μ Abs.:

56218



A= 0.0000e+000

B= 1.7711e-004

C= 2.6554e-002

Rho= 0.9999155

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.050	0.050	135	0.000	135				
0.2	0.200	0.208	0.008	1023	0.0 %	1023				
2.5	2.500	2.414	-0.086	13483	0.0 %	13483				
5.0	5.000	4.986	-0.014	28001	0.0 %	28001				
7.5	7.500	7.558	0.058	42527	0.0 %	42527				
10.0	10.000	9.983	-0.017	56218	0.0 %	56218				

LB112613 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Jan 2021 15:50:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	135	-	0.0000	7471B	PPB	04 Jan 2021 15:54:31	S	Std
0.2 - 1	50.2	1023	-	0.2000	7471B	PPB	04 Jan 2021 15:56:48	S	Std
2.5 - 1	52.5	13483	-	2.5000	7471B	PPB	04 Jan 2021 15:59:05	S	Std
5.0 - 1	55	28001	-	5.0000	7471B	PPB	04 Jan 2021 16:01:22	S	Std
7.5 - 1	57.5	42527	-	7.5000	7471B	PPB	04 Jan 2021 16:03:38	S	Std
10.0 - 1	60	56218	-	10.0000	7471B	PPB	04 Jan 2021 16:05:54	S	Std
ICV59 - 1	ICV59	21515	3.8370	-	7471B	PPB	04 Jan 2021 16:08:41	U	SMPL
ICB59 - 1	ICB59	64	0.0379	-	7471B	PPB	04 Jan 2021 16:10:56	U	SMPL
CCV65 - 1	CCV65	28434	5.0624	-	7471B	PPB	04 Jan 2021 16:13:11	U	SMPL
CCB65 - 1	CCB65	56	0.0365	-	7471B	PPB	04 Jan 2021 16:15:26	U	SMPL
CRA - 1	CRA	1012	0.2058	-	7471B	PPB	04 Jan 2021 16:17:41	U	SMPL
HighStd - 1	HighStd	56291	9.9961	-	7471B	PPB	04 Jan 2021 16:19:57	U	SMPL
ChkStd - 1	ChkStd	37418	6.6536	-	7471B	PPB	04 Jan 2021 16:22:12	U	SMPL
PB133903BL - 1	PB133903BL	-60	0.0159	-	7471B	PPB	04 Jan 2021 16:24:29	U	SMPL
PB133903BS - 1	PB133903BS	18689	3.3365	-	7471B	PPB	04 Jan 2021 16:26:45	U	SMPL
L5248-02 - 1	VR16125	8943	1.6104	-	7471B	PPB	04 Jan 2021 16:29:02	U	SMPL
L5249-01 - 1	BU-03-123120	1527	0.2970	-	7471B	PPB	04 Jan 2021 16:31:18	U	SMPL
L5249-01DUP - 1	BU-03-123120DUP	1368	0.2688	-	7471B	PPB	04 Jan 2021 16:33:35	U	SMPL
L5249-01MS - 1	BU-03-123120MS	20437	3.6461	-	7471B	PPB	04 Jan 2021 16:35:53	U	SMPL
L5249-01MSD - 1	BU-03-123120MSD	21497	3.8338	-	7471B	PPB	04 Jan 2021 16:38:08	U	SMPL
CCV66 - 1	CCV66	28075	4.9988	-	7471B	PPB	04 Jan 2021 16:40:23	U	SMPL
CCB66 - 1	CCB66	47	0.0349	-	7471B	PPB	04 Jan 2021 16:42:39	U	SMPL
L5249-01LX5 - 1		316	0.0825	-	7471B	PPB	04 Jan 2021 16:44:54	U	SMPL
L5249-01A - 1		24104	4.2955	-	7471B	PPB	04 Jan 2021 16:47:09	U	SMPL
CCV67 - 1	CCV67	28016	4.9884	-	7471B	PPB	04 Jan 2021 16:49:25	U	SMPL
CCB67 - 1	CCB67	45	0.0345	-	7471B	PPB	04 Jan 2021 16:51:41	U	SMPL