


LB112620

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9921e-004

C= -7.7377e-003

Rho= 0.9999930

Accept=Accepted

slope
Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.015	0.015	113	0.000	113					-
0.05	0.050	0.047	-0.003	277	0.0 %	277					-6
0.20	0.200	0.198	-0.002	1035	0.0 %	1035					-1
2.50	2.500	2.490	-0.010	12540	0.0 %	12540					0
5.00	5.000	5.006	0.006	25169	0.0 %	25169					0
7.50	7.500	7.475	-0.025	37560	0.0 %	37560					0
10.0	10.000	10.018	0.018	50330	0.0 %	50330					0

LB112620

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	113	-		0 SFAM01.1	PPB	1/5/2021 12:09	Std
	0.05 S01	277	-	0.05	SFAM01.1	PPB	1/5/2021 12:12	Std
	0.2 S02	1035	-	0.2	SFAM01.1	PPB	1/5/2021 12:14	Std
	2.5 S03	12540	-	2.5	SFAM01.1	PPB	1/5/2021 12:16	Std
	5 S04	25169	-	5	SFAM01.1	PPB	1/5/2021 12:18	Std
	7.5 S05	37560	-	7.5	SFAM01.1	PPB	1/5/2021 12:21	Std
	10 S06	50330	-	10	SFAM01.1	PPB	1/5/2021 12:25	Std
ICV004	ICV004	20335	4.0432	-	SFAM01.1	PPB	1/5/2021 12:28	SMPL
ICB004	ICB004	31	-0.0016	-	SFAM01.1	PPB	1/5/2021 12:30	SMPL
CCV068	CCV068	25501	5.0723	-	SFAM01.1	PPB	1/5/2021 12:33	SMPL
CCB068	CCB068	39	0	-	SFAM01.1	PPB	1/5/2021 12:35	SMPL
PB133923BL	PBS923	86	0.0094	-	SFAM01.1	PPB	1/5/2021 12:37	SMPL
L5214-03	MDL-SOIL-QT1-2021-01	266	0.0453	-	SFAM01.1	PPB	1/5/2021 12:42	SMPL
CCV069	CCV069	25325	5.0372	-	SFAM01.1	PPB	1/5/2021 12:44	SMPL
CCB069	CCB069	46	0.0014	-	SFAM01.1	PPB	1/5/2021 12:47	SMPL