

LB 112028


SFAM01.1

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

45137



A= 0.0000e+000

B= 2.2088e-004

C= -2.0453e-003

Rho= 0.9999501

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.015	0.015	79	0.000	79					14
0.05	0.050	0.057	0.007	268	0.0 %	268					-4
0.20	0.200	0.193	-0.007	884	0.0 %	884					-1
2.50	2.500	2.485	-0.015	11258	0.0 %	11258					-1
5.00	5.000	4.955	-0.045	22441	0.0 %	22441					1
7.50	7.500	7.577	0.077	34314	0.0 %	34314					0
10.0	10.000	9.968	-0.032	45137	0.0 %	45137					

LB112628
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	79	-		0 SFAM01.1	PPB	1/5/2021 13:49	Std
	0.05 S01	268	-	0.05	SFAM01.1	PPB	1/5/2021 13:51	Std
	0.2 S02	884	-	0.2	SFAM01.1	PPB	1/5/2021 13:53	Std
	2.5 S03	11258	-	2.5	SFAM01.1	PPB	1/5/2021 13:55	Std
	5 S04	22441	-	5	SFAM01.1	PPB	1/5/2021 13:58	Std
	7.5 S05	34314	-	7.5	SFAM01.1	PPB	1/5/2021 14:03	Std
	10 S06	45137	-	10	SFAM01.1	PPB	1/5/2021 14:08	Std
ICV005	ICV005	17865	3.944	-	SFAM01.1	PPB	1/5/2021 14:10	SMPL
ICB005	ICB005	-19	-0.0062	-	SFAM01.1	PPB	1/5/2021 14:13	SMPL
CCV070	CCV070	22073	4.8734	-	SFAM01.1	PPB	1/5/2021 14:15	SMPL
CCB070	CCB070	-36	-0.01	-	SFAM01.1	PPB	1/5/2021 14:17	SMPL
PB133922BL	PBW922	23	0.003	-	SFAM01.1	PPB	1/5/2021 14:19	SMPL
L5214-01	MDL-WATER-QT1-2021-01	233	0.0494	-	SFAM01.1	PPB	1/5/2021 14:22	SMPL
CCV071	CCV071	22111	4.8818	-	SFAM01.1	PPB	1/5/2021 14:24	SMPL
CCB071	CCB071	-22	-0.0069	-	SFAM01.1	PPB	1/5/2021 14:26	SMPL